

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040331_03

Certified AMS: O342M for O₃

Manufacturer: ENVEA
111, Boulevard Robespierre
78304 Poissy Cedex
France

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
VDI 4202-1 (2002), VDI 4203-3 (2004), EN 14625 (2012)
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 10 pages).
The present certificate replaces certificate 0000040331_02 dated 1 July 2020.



Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance
www.tuv.com
ID 0000040331


Publication in the German Federal Gazette
(BAnz) of 20 April 2007

German Environment Agency


Dessau, 27 June 2025

This certificate will expire on:
30 June 2030

TÜV Rheinland Energy &
Environment GmbH
Cologne, 26 June 2025



Dr. Marcel Langner
Head of Section II 4



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report:	936/21205818/B dated 8 December 2006 and Addendum 936/21221709/C dated 28 September 2013
Initial certification:	1 April 2014
Expiry date:	30 June 2030
Certificate:	Renewal (of previous certificate 0000040331_02 of 1 July 2020 valid until 30 June 2025)
Publication:	BAnz. 20 April 2007, No. 75, p. 4139, chapter III No. 3.1 and Banz AT 01.04.2014B12, chapter VI notification 21

Approved application

The tested AMS is suitable for continuous immission measurement of O₃ in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test.

The AMS is approved for an ambient temperature range of 0 °C to 30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

Basis of the certification

This certification is based on:

- Test report 936/21205818/B dated 8 December 2006 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH and Addendum 936/21221709/C dated 28 September 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 20 April 2007, No. 75, p. 4139, chapter III No. 3.1, Announcement by UBA dated 12 April 2007:

AMS designation:

O342M

Manufacturer:

Environnement S.A., Poissy Cedex, France and
Ansyco GmbH Karlsruhe, Germany

Field of application:

For continuous ambient air monitoring of ozone (stationary operation)

Measuring ranges during the performance test:

O ₃	0 – 360	µg/m ³
	0 – 500	µg/m ³

Software version:

V1.28

Test institute:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
TÜV Rheinland Group
Report No.: 936/21205818/B dated 8 December 2006

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, Chap. VI notification 21, Announcement by UBA dated 27 February 2014:

21 Notification of announcement by the German Federal Environment Agency of 12 April 2007 (BAnz. p. 4139, Chapter III Number 3.1)

The measuring system O342M for O₃, manufactured by Environnement, fulfils the requirements of Standard EN 14625 (December 2012). Furthermore, the manufacturing proves and the quality management system of the measuring system O342M for O₃ fulfil the requirements of EN 15267.

The test report of the performance test with report number 936/21205818/B as well as an addendum as an integral part of to the test report with report number 936/21221709/C can be viewed on the internet at www.qal1.de.

Statement by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, Chap. V notification 46,
Announcement by UBA dated 22 July 2015:

**46 Notification as regards Federal Environment Agency (UBA) notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 3.1) and
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter VI, notification 21)**

The current software version of the O342M measuring system for O₃,
manufactured by Environnement S.A., is:

v1.0.5 (Calculation process)

v3.6.b (Display process)

For additional means of communication, the AMS was equipped
with a USB and a TCP/IP interface.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 16 March 2015

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, Chap. IV
notification 31, Announcement by UBA dated 27 February 2019:

**31 Notification as regards Federal Environment Agency notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 3.1) and
of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V notification 46)**

The current software version of the O342M measuring system for ozone
manufactured by Environnement S.A. is:

v1.0.5 (calculation process)

v3.6.e (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV
notification 36, Announcement by UBA dated 24 February 2020:

**36 Notification as regards Federal Environment Agency (UBA) notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 3.1) and
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV notification 31)**

Environnement S.A., Poissy, France have changed their company name to ENVEA.

The latest software version of the O342M measuring system for ozone
manufactured by ENVEA is:

v1.0.5 (calculation process)

v3.8.b (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The O342M analyser measures ozone (O₃) in ambient air. It works by absorbing ozone in the UV range. The broad UV absorption spectrum of ozone has a maximum of approx. 254 nm. This wavelength precisely coincides with the strongest emission line of a mercury low-pressure lamp of 253.7 nm as well as with the maximum spectral sensitivity of the UV detector, a caesium telluride vacuum UV diode. This is an ideal requirement for photometric measurement of ozone.

This measuring principle corresponds to the standard reference method as described in EN 14625.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: gal1.de.

History of documents

Certification of O342M is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/21205818/B dated 8 December 2006
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Publication: BAnz. 20 April 2007, No. 75, p. 4139, chapter III number 3.1
UBA announcement dated 12 April 2007

Initial certification according to EN 15267 based on notification

Certificate No. 0000040331_00: 29 April 2014
Expiry date of the certificate: 31 March 2019
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013
Test report: 936/21205818/B dated 8 December 2006 and 936/21221709/C dated 28 September 2013
Publication: BAnz AT 01.04.2014 B12, chapter VI notification 21
UBA announcement dated 27 February 2014

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 16 March 2015
Publication: BAnz AT 26.08.2015 B4, chapter V notification 46
UBA announcement dated 22 July 2015
(Soft- and hardware changes)

Renewal of certificates

Certificate No. 0000040331_01: 1 April 2019
Expiry date of the certificate: 30 June 2020

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 31
UBA announcement dated 27 February 2019
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 36
UBA announcement dated 24 February 2020
(Software changes and new producer name formerly Environnement S.A.)

Renewal of certificates

Certificate No. 0000040331_02: 1 July 2020
Expiry date of the certificate: 30 June 2025

Renewal of certificates

Certificate No. 0000040331_03: 27 June 2025
Expiry date of the certificate: 30 June 2030

Expanded uncertainty laboratory, system 1

Measuring device:		Serial-No.:		Gerät 1	
Measured component:		1h-alert threshold:		120	
O ₃		1h-alert threshold:		120	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u _{r,z}	0.0025
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	1.000	u _{r,h}	0.0161
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-0.200	u _{l,h}	0.0192
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	-0.010	u _{gp}	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.023	u _{gt}	0.0705
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.035	u _{st}	0.0768
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.040	u _v	0.2904
8a	Interferent H ₂ O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	1.000	u _{H2O}	0.6737
		≤ 10 nmol/mol (Span)	1.100		
8b	Interferent Toluene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.100	u _{int, pos}	8.6700
		≤ 5.0 nmol/mol (Span)	2.600	or	
		≤ 5.0 nmol/mol (Zero)	0.200	u _{int, neg}	
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Span)	2.500		
9	Averaging effect	≤ 7.0% of measured value	3.000	u _{av}	4.3200
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{asc}	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{cg}	1.4400
Combined standard uncertainty			u _c	3.9482	
Expanded uncertainty			U	7.8964	
Relative expanded uncertainty			W	6.58	
Maximum allowed expanded uncertainty			W _{req}	15	

Expanded uncertainty laboratory, system 2

Measuring device:		Serial-No.:		Gerät 2	
Environment 0342M		1h-alert threshold:		120	
Measured component:		O ₃		nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u _{r,z}	0.0026
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0.800	u _{r,lv}	0.0108
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-1.300	u _{l,lv}	0.8112
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0.010	u _{gp}	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.040	u _{gt}	0.2133
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.065	u _{st}	0.2649
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.045	u _v	0.3675
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	0.000 0.500	u _{H2O}	0.1392
8b	Interferent Toluene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.100 1.500	u _{int,pos} or	4.0833
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.400 2.000	u _{int,neg}	
9	Averaging effect	≤ 7.0% of measured value	0.300	u _{av}	0.0432
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{dsc}	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{cg}	1.4400
Combined standard uncertainty				u _c	2.7175
Expanded uncertainty				U	5.4351
Relative expanded uncertainty				W	4.53
Maximum allowed expanded uncertainty				W _{req}	15

Combined uncertainty, laboratory and field, system 1

Measuring device:		Environnement O342M		Serial-No.:		Gerät 1	
Measured component:		O ₃		1h-alert threshold:		120	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	nmol/mol	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u _{r,z}	0.05	0.0025	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	1.000	u _{r,h}	not considered, as u _{r,h} = 0.12 < u _{r,f}	-	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-0.200	u _{l,h}	-0.14	0.0192	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	-0.010	u _{pp}	-0.09	0.0089	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.023	u _{gt}	0.27	0.0705	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.035	u _{st}	0.28	0.0768	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.040	u _v	-0.54	0.2904	
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	1.000	u _{H2O}	0.82	0.6737	
8b	Interferent Toluene with 0.5 µmol/mol	≤ 10 nmol/mol (Span)	1.100	u _{int,pos} or u _{int,neg}	2.94	8.6700	
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.100				
		≤ 5.0 nmol/mol (Span)	2.600				
9	Averaging effect	≤ 7.0% of measured value	3.000	u _{av}	2.08	4.3200	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.240	u _{r,f}	3.89	15.1165	
11	Long term drift at zero level	≤ 5.0 nmol/mol	-0.580	u _{d,l,z}	-0.33	0.1121	
12	Long term drift at span level	≤ 5.0% of max. of certification range	2.280	u _{d,l,h}	1.58	2.4952	
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{asc}	0.00	0.0000	
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{tg}	1.20	1.4400	
Combined standard uncertainty				u _c	5.7703	nmol/mol	
Expanded uncertainty				U	11.5405	nmol/mol	
Relative expanded uncertainty				W	9.62	%	
Maximum allowed expanded uncertainty				W _{req}	15	%	

Combined uncertainty, laboratory and field, system 2

Measuring device:			Environnement O342M		Serial-No.:		Gerät 2		nmol/mol
Measured component:			O ₃		1h-alert threshold:		120		
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty				
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u _{r,z}	0.05	0.0026			
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0.800	u _{r,h}	not considered, as u _{r,h} = 0.1 < u _{r,f}	-			
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-1.300	u _{l,h}	-0.90	0.8112			
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0.010	u _{dp}	0.09	0.0089			
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.040	u _{gt}	0.46	0.2133			
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.065	u _{st}	0.51	0.2649			
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.045	u _v	-0.61	0.3675			
8a	Interferent H ₂ O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0.000	u _{H2O}	0.37	0.1392			
8b	Interferent Toluene with 0.5 µmol/mol	≤ 10 nmol/mol (Span)	0.500	u _{tol,pos} or	2.02	4.0833			
		≤ 5.0 nmol/mol (Zero)	0.100						
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Span)	1.500	u _{tol,neg}	2.000				
		≤ 5.0 nmol/mol (Zero)	0.400						
9	Averaging effect	≤ 7.0% of measured value	0.300	u _{av}	0.21	0.0432			
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.240	u _{r,f}	3.89	15.1165			
11	Long term drift at zero level	≤ 5.0 nmol/mol	0.790	u _{d,z}	0.46	0.2080			
12	Long term drift at span level	≤ 5.0% of max. of certification range	3.550	u _{d,h}	2.46	6.0492			
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{asc}	0.00	0.0000			
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{tg}	1.20	1.4400			
			Combined standard uncertainty		u _c	5.3617			nmol/mol
			Expanded uncertainty		U	10.7234			nmol/mol
			Relative expanded uncertainty		W	8.94			%
			Maximum allowed expanded uncertainty		W _{req}	15			%