

# CONFIRMATION

Notification: 0000053805\_01\_01  
on changes according to EN 15267 regarding certificate 0000053805\_01 dated 02 March 2022

---

**AMS designation:** AC32e for NO, NO<sub>2</sub> and NO<sub>x</sub>

**Manufacturer:** ENVEA  
111, Boulevard Robespierre  
78304 Poissy Cedex  
France

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 09 March 2022  
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**10 Notification as regards Federal Environment Agency (UBA) notices  
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter III number 1.1) and  
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 19)**

The current software version for the measuring device AC 32e\*/AC 32e  
for NO, NO<sub>2</sub> and NO<sub>x</sub> from ENVEA is:

v1.1.c


The measuring device can be equipped alternatively with the photomultiplier of the  
company Hamamatsu of the type R5929.


Within the renewal of the certificate, new evaluations of measurement data were carried  
out. These are described in test report 936/21233023/C dated 7 January, 2022.

Statement issued by TÜV Rheinland Energy GmbH dated 07 January 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 12 April 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.