

## CONFIRMATION

Notification: 0000053805\_01\_02 on changes according to EN 15267 regarding certificate 0000053805\_01 dated 02 March 2022

AMS designation:

AC32e for NO, NO<sub>2</sub> and NO<sub>x</sub>

Manufacturer:

**FNVFA** 

111, Boulevard Robespierre

78304 Poissy Cedex

France

## **German Federal Environmental Agency (UBA)**

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6

- IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- Notification as regards Federal Environment Agency (UBA) notices of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter III number 1.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 10)

The current software version of the AC32e\*/AC32e measuring system for NO,  $NO_2$  and  $NO_x$  from ENVEA is:

v1.1.e

Furthermore the software version v1.1.d can be used.

The measuring system can alternatively be equipped with a revised high-voltage transformer for the ozone generator.

Statement issued by TÜV Rheinland Energy GmbH dated 04 January 2023

TÜV Rheinland Energy GmbH Cologne, 25 April 2023

i. V. Jol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu

TRE@umwelt-tuv.eu Tel. +49 221 806-5200 TÜV Rheinland Energy GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.