

## CONFIRMATION

Notification: 0000053805\_01\_03 on changes according to EN 15267 regarding certificate 0000053805\_01 dated 2 March 2022

AMS designation: AC32e for NO, NO<sub>2</sub> and NO<sub>x</sub>

Manufacturer: ENVEA

111, Boulevard Robespierre 78304 Poissy Cedex

France

## **German Federal Environmental Agency (UBA)**

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 2 April 2025
Publication: BAnz AT 19.05.2025 B3

- IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 89 Notification as regards Federal Environment Agency (UBA) notices of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter III number 1.1) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 64)

The current software version for the AC32e\*/AC32e measuring system for NO, NO $_{\rm 2}$  and NO $_{\rm x}$  from ENVEA is

v1.2.c

The software versions v1.2.a and v1.2.b can also be used.

The measuring system can be equipped with a revised version of the ARM20 board.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 04 October 2024

TÜV Rheinland Energy & Environment GmbH Cologne, 20 May 2025

i. V. Mol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu TRE@umwelt-tuv.eu

Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.