

## CONFIRMATION

Notification: 0000074620\_00\_02 on changes according to EN 15267 regarding certificate 0000074620\_00 dated 17 June 2020

AMS designation:

MP101M for PM<sub>10</sub>

Manufacturer:

**ENVEA** 

111, Boulevard Robespierre

78304 Poissy Cedex

France

## **German Federal Environmental Agency (UBA)**

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4

- III. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 40 Notification as regards Federal Environment Agency (UBA) notices of 31 March 2020 (BAnz AT 07.05.2020 B8, chapter II number 2.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 8)

The current software of the measuring device MP101M for suspended particulate matter  $PM_{10}$  of the company ENVEA reads:

v.4.0.I

The module board in the measuring device can also be used in revision H.

Statement issued by TÜV Rheinland Energy GmbH dated 28 April 2022

TÜV Rheinland Energy GmbH Cologne, 01 August 2022

i. V. Mil.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu

TRE@umwelt-tuv.eu Tel. +49 221 806-5200 TÜV Rheinland Energy GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.