

CONFIRMATION

Notification: 0000072200_00_07
on changes according to EN 15267 regarding certificate 0000072200_00 dated 4 June 2020

AMS designation: CMM AutoQAL for Hg

Manufacturer: Gasmot Technologies Oy
Mestarintie 6
01730 Vantaa
Finland

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BA_nz AT 10.05.2024 B7**

- V. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**
- 24 Notification as regards Federal Environment Agency (UBA) notices
of 24 February 2020 (BA_nz AT 24.03.2020 B7, chapter I number 1.1) and
of 21 February 2023 (BA_nz AT 20.03.2023 B6, chapter IV notification 29)**


The current software version for the CMM AutoQAL measuring system for the Hg component from the manufacturer Gasmot Technology Oy is:


1.2080.

Furthermore, the alternative vaporiser nozzle type ARG-1-US6 can be used in the integrated test gas generator.

Statement issued by TÜV Rheinland Energy GmbH dated 15 December 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.