

CERTIFICATE

about Product Conformity (QAL1)

Number of Certificate: 0000028730

Certified AMS: GIGAS 10M for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂

Manufacturer: General Impianti S.r.l.
Via Monteschiario 3
60030 Moie di Maiolati
Italy

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

This is certifying that the AMS has been tested
and found to comply with:

EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007
and EN 14181: 2004

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).




- EN 15267-3 tested
- QAL1 certified
- TÜV approved
- Annual inspection


Publication in the German Federal Gazette
(BAnz.) of 29 July 2011

The certificate is valid until:
28 July 2016

Umweltbundesamt
Dessau, 19 August 2011

TÜV Rheinland Energie und Umwelt GmbH
Köln, 17 August 2011


i. A. Dr. Hans-Joachim Hummel


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 - 221 - 806 - 2756

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Köln

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

Test report:	936/21211855/B of 25 March 2011
First certification:	29 July 2011
Run of validity until:	28 July 2011
Publication	BAnz. 29 July 2011, No 113, page 2725, chapter I No 4.3

Authorised application

Suitability of the AMS for application at plants requiring licensing and plants according to the 27 BImSchV was assessed on the basis of a laboratory test and a field test over a more than 15 month period of the GIGAS 10M measuring system for monitoring the components HF and N₂O at a tunnel kiln plant for firing refractory and acid proof bricks.

The AMS is approved for the temperature range from +5 °C to +40 °C.

The measuring system had been certified and approved on the basis of extensive laboratory tests and a field test of more than 12 months at a municipal waste incineration plant for measuring the components CO, NO, NO₂, SO₂, HCl, NH₃, CO₂ and H₂O during an earlier suitability test. The tested measurement ranges were selected in order to secure an application range for the AMS as wide as possible.

The results of the earlier tested procedure for CO, NO, NO₂, SO₂, HCl, NH₃, CO₂ and H₂O as well as the test results of the current testing for HF and N₂O have been assessed and found to comply with the requirements of the latest European Standard for pollution control purposes (QAL1 according to EN 15267).

Any potential user should ensure in consultation with the manufacturer that this AMS is suitable for the facility on which it will be installed.

Basis of the certification

This certification is based on:

- the test report 936/21211855/B of 25 March 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Environmental Agency (UBA) as relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 29 July 2011, No 113, p. 2725, chapter I No 4.3: UBA publication from 15 July 2011)

AMS name:

GIGAS 10M for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂

Manufacturer:

General Impianti S.r.l., Moie di Maiolati, Italy

Suitability:

For measurements at plants requiring official permission (i. e. plants in 2000-76-EC, waste incineration directive and 2001-80-EC, large combustion plants directive).

Measurement ranges during the suitability test:

Component	Certification range	Supplementary measurement ranges	Unit
HF	0 - 5	0 - 10 0 - 20	mg/m ³
N ₂ O	0 - 50	0 - 1000	mg/m ³
CO	0 - 75	0 - 300	mg/m ³
SO ₂	0 - 75	0 - 300	mg/m ³
NO	0 - 200	0 - 400	mg/m ³
NO ₂	0 - 100	0 - 200	mg/m ³
HCl	0 - 15	0 - 90	mg/m ³
NH ₃	0 - 15	-	mg/m ³
CO ₂	0 - 20	-	Vol.-%
H ₂ O	0 - 30	-	Vol.-%

Software versions: Omnic 7.2
GasCalc: 4.4

Restriction:

The measurement system shall only be operated at plants waste gas humidity does not constantly exceed 30 Vol.-%.

Remarks:

1. Wet test gases shall be used for the testing of HF, HCl, and NH₃.
2. A six month period has been determined as maintenance interval.
3. Supplementary testing (including the components N₂O and HF, instrument changes and conversion of test results to standard EN 15267-3) on the announcements of the Federal Environment Agency on 12 August 2008 (BAZ. p. 3243, chapter I No 2.3).
4. For the measuring component CO the requirement for the total uncertainty according to EN 15267-3 is not fulfilled.
5. The measuring unit works with wet process gases.

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Köln
Report-No: 936/21211855/B of 25 March 2011

Certified product

This certificate applies to automatic measurement systems that comply with the following description:

The GIGAS 10M measuring system is an extractive multiple-component measuring system based on the measuring principle of FTIR spectrometry which measures at high temperatures. It comprises the main components as described below:

Sampling

Sampling probe: General Impianti GL – SRPF (180 °C) coated
Sampling tube: RACO (Length during suitability testing approx. 10 m,
heated to 180 °C)
Heated filter: M&C – FT-H2 (180 °C)

Analyser

FTIR: GIGAS 10M , temperature of the cuvette: 180 °C

Sample gas post-treatment

The following components are installed after the sample gas outlet:

Sample gas cooler: General Impianti – FRIGO GI PELLTIER R
Sample gas pump: KNF – N.814.KTE

Control modules

DAQ module: GL-AnDe
Omron module: GL-TPReg

Calculator

Standard PC of the following minimum requirements:

Operating system: MS Windows XP
Processor: Intel Pentium III, 1 GHz
Primary storage: 512 MB
Hard disk: 40 GB
Interfaces: USB Interface
Network interface RJ 45
Serial Interface RS 232

A Siemens Industry PC with 17" Touch Screen Display has been installed during the suitability test.

Software

Evaluation-Software: GasCalc 4.4 and Omnic 7.2

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for a long-lasting compliance of the ongoing production process with the requirements of EN 15267. The manufacturer is obliged to maintain a certified quality management system to control the production of the certified product. Both product and quality management system shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the given address on page 1.

The certification mark with the product specific ID-Number which may be applied to the product or used in promotion material of the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remain property of TÜV Rheinland Energie und Umwelt GmbH. Upon revocation of the announcement the certificate loses validity. After expiration of the validity of the certificate or on request of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certification mark shall longer be used.

The current version of this certificate and its validity is also listed at the Internet Address: qal1.de.

Certification of GIGAS 10M for HF, N₂O, CO, NO, NO₂, SO₂, HCl, NH₃, H₂O and CO₂ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/21206517/A from 08 July 2007
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Köln

Publication: BAnz. 06 November 2007, No 206, p. 7925, chapter I No 2.1:
UBA announcement from 23 September 2007

Test report: 936/21206517/B from 09 November 2007
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Köln

Publication: BAnz. 07 March 2008, No 38, p. 901, chapter I No 2.3:
UBA announcement from 14 February 2008

Test report: 936/21206517/C from 27 February 2008
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Köln

Publication: BAnz. 03 September 2008, No 133, p. 3242, chapter I No 2.3:
UBA announcement from 12 August 2008

Notification

Publication: BAnz. 26 January 2011, No 14, p. 294, chapter IV notification 29:
UBA announcement from 10 January 2011 (Software)

Initial certification according to EN 15267

Certificate No 0000028730 of: 19 August 2011

Validity of the certificate: 28 July 2011

Test report: 936/21211855/B of 25 March 2011
TÜV Rheinland Energie und Umwelt GmbH, Köln

Publication: BAnz. 29 July 2011, No 113, p. 2725, chapter I No 4.3:
UBA publication of 15 July 2011

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	RSE09/TUV/H1 / RSE09/TUV/H2
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	HF 0 - 5 mg/m ³
---------------------	-------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.20 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	0.15 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	0.20 mg/m ³
Uncertainty of cross sensitivity	0.12 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.080 mg/m ³	0.006 (mg/m ³) ²
Lack of fit	u _{lof} -0.052 mg/m ³	0.003 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.066 mg/m ³	0.004 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.084 mg/m ³	0.007 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.051 mg/m ³	0.003 (mg/m ³) ²
Influence of supply voltage	u _v 0.029 mg/m ³	0.001 (mg/m ³) ²
Cross sensitivity (interference)	u _i 0.115 mg/m ³	0.013 (mg/m ³) ²
Influence of sample gas flow	u _p 0.046 mg/m ³	0.002 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.040 mg/m ³	0.002 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 0.20 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 0.40 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 2 mg/m³ 19.9

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 2 mg/m³ 40.0

Requirement of EN 15267-3

U in % of the ELV 2 mg/m³ 30.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	RSE09/TUV/H1 / RSE09/TUV/H2
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	N ₂ O 0 - 50 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.31 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	0.93 mg/m ³
Sum of negative CS at reference point	-1.98 mg/m ³
Maximum sum of cross sensitivities	-1.98 mg/m ³
Uncertainty of cross sensitivity	-1.14 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.100 mg/m ³	0.010 (mg/m ³) ²
Lack of fit	u _{lof} 0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.231 mg/m ³	0.053 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.808 mg/m ³	0.653 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.321 mg/m ³	0.103 (mg/m ³) ²
Influence of supply voltage	u _v 0.128 mg/m ³	0.016 (mg/m ³) ²
Cross sensitivity (interference)	u _i -1.143 mg/m ³	1.307 (mg/m ³) ²
Influence of sample gas flow	u _p 0.264 mg/m ³	0.070 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	1.56 mg/m ³
Total expanded uncertainty	U = u _c * k = u _c * 1.96	3.05 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 20 mg/m³	15.3
U in % of the ELV 20 mg/m³	20.0
U in % of the ELV 20 mg/m ³	15.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	CO 0 - 75 mg/m ³
---------------------	--------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.41 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	3.00 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.00 mg/m ³
Uncertainty of cross sensitivity	1.732 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.407 mg/m ³	0.166 (mg/m ³) ²
Lack of fit	u _{lof} -0.404 mg/m ³	0.163 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.476 mg/m ³	0.227 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.996 mg/m ³	0.992 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.321 mg/m ³	0.103 (mg/m ³) ²
Influence of supply voltage	u _v 0.093 mg/m ³	0.009 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.732 mg/m ³	3.000 (mg/m ³) ²
Influence of sample gas flow	u _p 0.433 mg/m ³	0.187 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	2.28 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	4.48 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³	9.0
U in % of the ELV 50 mg/m³	10.0
U in % of the ELV 50 mg/m ³	7.5

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NO 0 - 200 mg/m ³
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	-6.80 mg/m ³
Sum of positive CS at reference point	2.60 mg/m ³
Sum of negative CS at reference point	-5.20 mg/m ³
Maximum sum of cross sensitivities	-6.80 mg/m ³
Uncertainty of cross sensitivity	-3.926 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.782 mg/m ³	3.176 (mg/m ³) ²
Lack of fit	u _{lof} 1.155 mg/m ³	1.334 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.808 mg/m ³	0.653 (mg/m ³) ²
Span drift from field test	u _{d,s} -3.002 mg/m ³	9.012 (mg/m ³) ²
Influence of ambient temperature at span	u _t 1.650 mg/m ³	2.723 (mg/m ³) ²
Influence of supply voltage	u _v 0.513 mg/m ³	0.263 (mg/m ³) ²
Cross sensitivity (interference)	u _i -3.926 mg/m ³	15.413 (mg/m ³) ²
Influence of sample gas flow	u _p 1.155 mg/m ³	1.334 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 1.617 mg/m ³	2.613 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 6.04 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 11.84 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 130.4 mg/m³ 9.1

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 130.4 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 130.4 mg/m³ 15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NO ₂ 0 - 100 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	3.99 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	3.50 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.99 mg/m ³
Uncertainty of cross sensitivity	2.304 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.864 mg/m ³	0.746 (mg/m ³) ²
Lack of fit	u _{lof} 0.924 mg/m ³	0.854 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.346 mg/m ³	0.120 (mg/m ³) ²
Span drift from field test	u _{d,s} -1.559 mg/m ³	2.430 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.306 mg/m ³	0.094 (mg/m ³) ²
Influence of supply voltage	u _v 0.289 mg/m ³	0.084 (mg/m ³) ²
Cross sensitivity (interference)	u _i 2.304 mg/m ³	5.307 (mg/m ³) ²
Influence of sample gas flow	u _p 0.577 mg/m ³	0.333 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.808 mg/m ³	0.653 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 3.26 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 6.39 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 60 mg/m³ 10.6

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 60 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 60 mg/m³ 15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	SO ₂ 0 - 75 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.89 mg/m ³
Sum of negative CS at zero point	-0.53 mg/m ³
Sum of positive CS at reference point	3.00 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	3.00 mg/m ³
Uncertainty of cross sensitivity	1.732 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.263 mg/m ³	0.069 (mg/m ³) ²
Lack of fit	u _{lof} -0.572 mg/m ³	0.327 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.563 mg/m ³	0.317 (mg/m ³) ²
Span drift from field test	u _{d,s} 1.212 mg/m ³	1.469 (mg/m ³) ²
Influence of ambient temperature at span	u _t 1.664 mg/m ³	2.769 (mg/m ³) ²
Influence of supply voltage	u _v 0.179 mg/m ³	0.032 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.732 mg/m ³	3.000 (mg/m ³) ²
Influence of sample gas flow	u _p 0.433 mg/m ³	0.187 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 2.92 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 5.73 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 50 mg/m³ 11.5

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 50 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 15.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	HCl 0 - 15 mg/m ³
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.49 mg/m ³
Sum of negative CS at zero point	-0.61 mg/m ³
Sum of positive CS at reference point	0.60 mg/m ³
Sum of negative CS at reference point	-0.15 mg/m ³
Maximum sum of cross sensitivities	-0.61 mg/m ³
Uncertainty of cross sensitivity	-0.350 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.144 mg/m ³	0.021 (mg/m ³) ²
Lack of fit	u _{lof} -0.104 mg/m ³	0.011 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.251 mg/m ³	0.063 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.251 mg/m ³	0.063 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.186 mg/m ³	0.035 (mg/m ³) ²
Influence of supply voltage	u _v 0.026 mg/m ³	0.001 (mg/m ³) ²
Cross sensitivity (interference)	u _i -0.350 mg/m ³	0.122 (mg/m ³) ²
Influence of sample gas flow	u _p 0.087 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.121 mg/m ³	0.015 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.58 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.14 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 10 mg/m³	11.4
U in % of the ELV 10 mg/m³	40.0
U in % of the ELV 10 mg/m³	30.0

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	NH ₃ 0 - 15 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.52 mg/m ³
Sum of negative CS at zero point	-0.27 mg/m ³
Sum of positive CS at reference point	0.60 mg/m ³
Sum of negative CS at reference point	-0.15 mg/m ³
Maximum sum of cross sensitivities	0.60 mg/m ³
Uncertainty of cross sensitivity	0.346 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.086 mg/m ³	0.007 (mg/m ³) ²
Lack of fit	u _{lof} 0.165 mg/m ³	0.027 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.147 mg/m ³	0.022 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.251 mg/m ³	0.063 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.173 mg/m ³	0.030 (mg/m ³) ²
Influence of supply voltage	u _v 0.017 mg/m ³	0.000 (mg/m ³) ²
Cross sensitivity (interference)	u _i 0.346 mg/m ³	0.120 (mg/m ³) ²
Influence of sample gas flow	u _p 0.087 mg/m ³	0.008 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.121 mg/m ³	0.015 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 0.54 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 1.06 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 10 mg/m³	10.6
U in % of the ELV 10 mg/m³	40.0
U in % of the ELV 10 mg/m ³	30.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

The chosen value is recommended by the certification body.

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	CO ₂ 0 - 20 Vol.-%
---------------------	----------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.00 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.00 Vol.-%
Uncertainty of cross sensitivity	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.067 Vol.-%	0.004 (Vol.-%) ²
Lack of fit	u _{lof} -0.104 Vol.-%	0.011 (Vol.-%) ²
Zero drift from field test	u _{d,z} -0.058 Vol.-%	0.003 (Vol.-%) ²
Span drift from field test	u _{d,s} -0.231 Vol.-%	0.053 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.252 Vol.-%	0.064 (Vol.-%) ²
Influence of supply voltage	u _v 0.026 Vol.-%	0.001 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.000 Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _p 0.115 Vol.-%	0.013 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.162 Vol.-%	0.026 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.42 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.82 Vol.-%

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the range 20 Vol.-%	4.1
U in % of the range 20 Vol.-%	10.0
U in % of the range 20 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

The chosen value is recommended by the certification body.

*** and RSE09/TUV/H1 / RSE09/TUV/H2

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	General Impianti
Name of measuring system	GIGAS 10M
Serial number of the candidates	S1 A210015 / S2 A20016 ***
Measuring principle	FTIR

Test report

Test laboratory	936/21211855/B TÜV Rheinland
Date of report	2011-03-25

Measured component

Certification range	H ₂ O 0 - 30 Vol.-%
---------------------	-----------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.00 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.00 Vol.-%
Uncertainty of cross sensitivity	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.208 Vol.-%	0.043 (Vol.-%) ²
Lack of fit	u _{lof} -0.173 Vol.-%	0.030 (Vol.-%) ²
Zero drift from field test	u _{d,z} -0.017 Vol.-%	0.000 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.468 Vol.-%	0.219 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.172 Vol.-%	0.030 (Vol.-%) ²
Influence of supply voltage	u _v 0.015 Vol.-%	0.000 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.000 Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _p 0.173 Vol.-%	0.030 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.242 Vol.-%	0.059 (Vol.-%) ²

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.64 Vol.-%
Total expanded uncertainty	U = u _c * k = u _c * 1.96	1.26 Vol.-%

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the range 30 Vol.-%	4.2
U in % of the range 30 Vol.-%	10.0
U in % of the range 30 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.
The chosen value is recommended by the certification body.

*** and RSE09/TUV/H1 / RSE09/TUV/H2