



Confirmation

Notification: 4061336-2-ts

on changes in accordance with DIN EN 15267 regarding certificate no. 4061336-ts from 24th June 2025

AMS	IR-Pyrometer KT 15.67 IIP
Manufacturer	Heitronics Infrarot Messtechnik GmbH Kreuzberger Ring 40 65205 Wiesbaden Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and immissions
dated 10th July 2026

Publication: BAnz AT 10.07.2026 B5

VI. Notifications on Uniform nationwide practice for monitoring emissions and immissions

8 Notification on the UBA announcements from 2nd April 2025 (BAnz AT 19.05.2025 B3, Chapter III Number 1.1) and from 27th August 2025 (BAnz AT 31.10.2025 B5, Chapter IV 32. Notification)

The current software version of the AMS KT 15.67 IIP for monitoring minimum temperature in approved plants and plants according to 27th BImSchV by Heitronics Infrarot Messtechnik GmbH, Wiesbaden, is 5.19.

Version 5.14 can also be used.

The AMS KT 15.67 IIP for monitoring minimum temperature in approved plants and plants according to 27th BImSchV by Heitronics Infrarot Messtechnik GmbH, Wiesbaden, can also be optionally equipped with the protective housing WK15.

Comment of TÜV Süd Industrie Service GmbH from 19th October 2025.

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/ calibration
Munich, 09.09.2026


M.Sc. Simon Gerstlacher


Dipl.-Ing. Univ. Lukas Zielbauer