

Confirmation

Notification: 4139485_01
on design changes in accordance with DIN EN 15267 regarding certificate no. 3919266-ts from 23 May 2024

AMS	IR-Pyrometer KT 15.69 IIP
Manufacturer	Heitronics Infrarot Messtechnik GmbH Kreuzberger Ring 40 65205 Wiesbaden Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in monitoring emissions and immissions
dated 2 April 2025
Publication: BAnz AT 19.05.2025 B3**

IV. Notification on uniform practise

41 Notification on the UBA announcement from 19 March 2024 (BAnz AT 10.05.2024 B7, Chapter II No. 1.2)

As an alternative to the B2-Li flange, the AMS KT 15.69 IIP can be mounted with the B2 flange at the measuring point.

As an alternative to the BX Kopf adapter, BX M72 adapter and B4 M72 adapter the AMS KT 15.69 IIP can be mounted with the B4 and B7 adapter at the measuring point.

Comment of TÜV Süd Industrie Service GmbH dated 29 August 2024

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/ calibration
Munich, 30 April 2025



Dipl.-Ing. Hans-Jörg Eisenberger



M.Sc. Ulrich Peschke