

Confirmation

Notification: 4149451_01
on design changes in accordance with DIN EN 15267 regarding certificate no. 3919266-ts from 23 May 2024

AMS IR-Pyrometer KT 15.69 IIP

Manufacturer Heitronics Infrarot Messtechnik GmbH
Kreuzberger Ring 40
65205 Wiesbaden
Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and immissions dated 27 August 2025
Publication: AT 31.10.2025 B5

IV. Notification on uniform practise

30 Notifications on the UBA announcement from 19 March 2024 (BAnz AT 10.05.2024 B7, Chapter II No. 1.2) and from 2 April 2025 (BAnz AT 19.05.2025 B3, Chapter IV No. 41)

The measuring system KT 15.69 IIP for monitoring the minimum temperature in plants requiring authorisation and plants in compliance with the 27th BImSchV, manufactured by Heitronics Infrarot Messtechnik GmbH, Wiesbaden, can alternatively be operated with the following 24 V DC power supplies for power supply:

1. Siemens LOGO!Power® power supply unit 24 V/0.6 A (Article number 6EP3330-6SB00-0AY0) for parallel power supply of up to 2 measuring systems,
2. Siemens SITOP® power supply unit 24 V/5 A (Article number 6EP3333-7SB00-0AX0) for parallel power supply of up to 16 measuring systems.

The power supply units are intended for installation in control cabinets.

Comment of TÜV Süd Industrie Service GmbH dated 19 February 2025

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/ calibration
Munich, 14 November 2025



Dipl.-Ing. Hans-Jörg Eisenberger



M.Sc. Ulrich Peschke