

CERTIFICATE

of product conformity (QAL 1)

Certificate number: 3533424-ts

Certified AMS	KT 19.67 II for temperature measurement in combustion gases
Manufacturer	Heitronics Infrarot Messtechnik GmbH Kreuzberger Ring 40 65205 Wiesbaden Germany

Test institute TÜV SÜD Industrie Service GmbH

This is to certify that the AMS has been tested and found to comply with the standards
DIN EN 15267-1 (2009), DIN EN 15267-2 (2023), DIN EN 15267-3 (2008) and
DIN EN 14181 (2015).

Certification applies to the conditions listed in this certificate
(the certificate consists of 6 pages).



Certificate No.: 3533424-ts

Publication in the German Federal Gazette
(BAnz) of 19 May 2025

This certificate will expire on
18 May 2030

Umweltbundesamt
Dessau, 24 June 2025

TÜV SÜD Industrie Service GmbH
Testing laboratory emission measurement/
calibration
Munich, 23 June 2025



Dr. Marcel Langner
Head of Section II 4



Hans-Jörg Eisenberger

Test report	3533424_V2 from 29 September 2024
Initial certification	19 May 2025
Certification validity until	18 May 2030 (5 years)
Publication	BAnz AT 19 May 2025 B3, chapter III, no. 1.2

Approved application

The tested AMS is suitable for use at plants requiring authorisation (13. BImSchV:2021, 17. BImSchV:2024) as well as at plants in accordance with the 27. BImSchV:2013. The suitability for this applications were assessed on the basis of a laboratory test and a shortened field test over six weeks of the AMS in the afterburning zone at plant according to Directive 2010/75/EU chapter IV (17. BImSchV). The measuring system is approved for ambient temperatures between -20 °C and +50 °C.

The AMS publication, the suitability test and the performance of the uncertainty calculations were conducted based on the provisions valid at the time of testing. Due to possible amendments to legal foundations, every user should ensure before use of the AMS that it is suitable for monitoring the applicable values.

The operator should consult the manufacturer to ensure that the AMS is suitable for the plant at which it is to be installed.

Note:

The legal regulations mentioned do not always have to correspond to the current state of legislation. Each user should ensure, if necessary in consultation with the competent authority, that this AMS fulfils the legal requirements for the intended use. Furthermore, it cannot be ruled out that legal regulations on the use of a measuring system for emission monitoring may change during the term of the certificate.

Certification basis

This certificate is based on:

- TÜV SÜD Industrie Service GmbH test report 3533424_V2 from 29 September 2024
- Suitability announcement by the German Federal Environmental Agency as relevant body
- The ongoing surveillance of the product and the manufacturing process

- Publication in the German Federal Gazette (BAnz AT 19 May 2025 B3, chapter III, no. 1.2, UBA publication from 2 April 2025):

AMS: IR-Pyrometer KT 19.67 II

Manufacturer: Heitronics GmbH Infrarot Messtechnik, Wiesbaden

Suitability: For plants requiring authorisation in compliance with the 13 BImSchV and the 17. BImSchV and for plants in compliance with the 27. BImSchV for monitoring the minimum temperature in combustion gases

Measurement range in the suitability test:

Component	Certification range	Unit
Temperature	400 - 1400	°C

Software version: 5.14

Restrictions:

None

Notes:

1. The maintenance interval is four weeks.
2. The AMS must be verified annually by using a Planckian radiator.
3. The AMS is suitable to detect gas temperature up to a maximum penetration depth of 4 metres.

Test report: TÜV Süd Industrie Service GmbH, Munich
Report no.: 3533424_V2 from 29 September 2024

Certified Product

The certificate applies to AMS that comply with the following description:

The tested measuring system consists of the components for the radiation pyrometer with a fixed focused long lens, an optical sight, evaluation and operating software as well as optionally with a shockblower.

The radiation pyrometer KT 19.67 II works in a spectral range where hot Carbondioxide over 400 °C has a large emissivity. However cold CO₂ is largely transmissive. For a selective filtering to a specific wavelength (4,58 µm) an interference filter is used. Because of the seal-air tube with schockblower the pyrometer is protected from dust and corrosive gases. Optionally a shockblower can be used.

The entire system consists of the following components:

Radiation pyrometer

Manufacturer: Heitronics Infrarot Messtechnik GmbH, 65205 Wiesbaden
Type: KT 19.67 II

Components

Power supply unit

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: T24 II

Adapter

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: BX tube with connection for seal-air

Adapter

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: B4 M72

Adapter

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: BX M72 with Sapphire disk

Adapter

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: B2-Li

Adjustable flange

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: B5J

Adapter (optional)

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: B4

Adapter (optional)

Manufacturer: Heitronics Infrarot Messtechnik GmbH
Type: B7

Flansch (optional)

Manufacturer:

Heitronics Infrarot Messtechnik GmbH

Type:

B2

Schockblower (option)

Manufacturer:

VSR Industrietechnik GmbH, 47189 Duisburg

Type:

VSR Blaster air injector

General notes

This certificate is based on the analyser tested. The manufacturer is responsible for the continuous compliance of the production to the DIN EN 15267 requirements. The manufacturer is required to maintain an approved quality management system to control the manufacture of the certified product. Regular monitoring must be conducted on both the product and the quality management systems.

If the product from the current production series no longer comply with the certified product, the Environmental Service Department of TÜV SÜD Industrie Service GmbH must be informed (address see footnote).

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied on the product or used in publicity material for the certified product.

This document and the certification mark shall remain the property of TÜV SÜD Industrie Service GmbH.

Should the publication be revoked, this certificate will become invalid. This document must be returned when the period of validity has elapsed and at the request of TÜV SÜD Industrie Service GmbH and the certification mark may no longer be used.

The current version of the certificate and its expiration is also accessible on the internet at **qal1.de**.

The certification of KT 19.67 II is based on the following documents and the regular continuous monitoring of the manufacturer's quality management system:

Initial certification in accordance with DIN EN 15267:

Certificate no.: 353424-ts

19 May 2025

Certificate validity until

18 May 2030 (5 years)

Report no.: 3533424_V2 from 29 September 2024

TÜV SÜD Industrie Service GmbH

Publication: BAnz AT 19 May 2025 B3, chapter III no. 1.2

UBA publication from 02 April 2025

Calculation of total uncertainty for QAL1 testing to DIN EN 14181 and DIN EN 15267-3 for KT 19.67 II

Total uncertainty for measurement range 400 – 1400 °C

Performance characteristic	Uncertainty	Value standard uncertainty °C	Square of standard uncertainty °C²
Lack-of-fit	u_{lof}	-1,155	1,334
Zero drift	$u_{d,z}$	-9,238	85,3406
Span drift	$u_{d,s}$	-16,166	261,3396
Influence of ambient temperature at span	u_t	2,410	5,8081
Influence of sample gas pressure	u_p		
Influence of sample gas flow	u_f		
Influence of supply voltage	u_v	0,242	0,0586
Cross-sensitivity (interferences)	u_i		
Repeatability standard deviation at reference point	$u_r = s_r$	0,061	$u_r < du$
Standard deviation from paired measurements	$u_d = s_d$	9,175	84,1806
Uncertainty of transfer-standard-pyrometers 5,4°C	u_{rm}	5,4	29,16
Excursion of measurement beam	u_{mb}		
Converter efficiency for AMS measuring NOx	u_{ce}		
Variation of response factors (TOC)	u_{rf}		
		Sum	467,2215
Combined standard uncertainty	$u_c = \sqrt{\sum(u_i)^2}$	21,6153	°C
Total expanded uncertainty	$U_{0,95} = 1,96 \times u_c$	42,366	°C
Relativ expanded uncertainty	U	5,0	% ELV
Permissible uncertainty of EN 15267-3	(at ELV 850 °C)	7,5	% ELV
Complied with requirements relating to the measurement uncertainty		yes	regarding EN 15267-3
Permissible uncertainty 13. / 17. BImSchV	(at ELV 850 °C)	10	% ELV
Complied with requirements relating to the measurement uncertainty		yes	regarding 13. / 17. BImSchV