

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000028754_05

Certified AMS: APMA 370 for CO

Manufacturer: Horiba Ltd.
2 Miyanoigashi
Kisshoin Minami-ku / Kyoto 610-8510
Japan

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
VDI 4202-1 (2010), VDI 4203-3 (2010), EN 14626 (2012),
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 15 pages).

The present certificate replaces certificate 0000028754_04 dated 25 January 2021.




Publication in the German Federal Gazette
(BAnz) of 8 April 2006


This certificate will expire on:
25 January 2031

German Environment Agency

TÜV Rheinland
Energy & Environment GmbH
Cologne, 21 January 2026

Dessau, 22 January 2026


Dr. Marcel Langner
Head of Section II 4


ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu qal1-info@tuv.com Tel. + 49 221 806-5200	TÜV Rheinland Energy & Environment GmbH Am Grauen Stein 51105 Köln
Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.	

Test report: 936/21204643/B dated 5 January 2006

Initial certification: 26 January 2011

Expiry date: 25 January 2031

Certificate: Renewal (of previous certificate 0000028754_04 of
25 January 2021 valid until 25 January 2026)

Publication: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV No. 2.1

Approved application

The tested AMS is suitable for continuous immission measurement of CO in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test.

The AMS is approved for an ambient temperature range of 0 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

Basis of the certification

This certification is based on:

- Test report 936/21204643/B dated 5 January 2006 of TÜV Immissionsschutz und Energiesysteme GmbH
- Addenda 936/21204643/B1 dated 27 July 2011 and 936/21222689/B dated 5 October 2013
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV No. 2.1, Announcement by UBA dated 21 February 2006:

AMS designation:

APMA 370

Manufacturer:

HORIBA Ltd., Kyoto 610 - 8510, Japan

Distribution: HORIBA Europe GmbH, 42799 Leichlingen

Field of application:

For continuous ambient air monitoring of CO (stationary operation)

Measuring ranges during the performance test:

Component	Certification range	Supplementary measurement ranges	Unit
CO	0 – 60	0 – 100	mg/m ³

Software version:

Version P1000878001C

Test institute:

TÜV Immissionsschutz und Energiesysteme GmbH, Cologne

TÜV Rheinland Group

Report No.:

936/21204643/B dated 5 January 2006

Publication in the German Federal Gazette: BAnz. 25. August 2009, Nr. 125, S. 2929, Chap. III notification 1, Announcement by UBA dated 3 August 2009:

1 Notification as regards Federal Environment Agency notice of 21 February 2006 (BAnz. p. 2655)

The latest software version of the APMA 370 ambient air measuring system manufactured by Horiba Europe GmbH is:
P1000878001J

The type GD-6 EH sample gas pump manufactured by Horiba may be used instead of the N 86 KNE sample gas pump manufactured by KNF.

Statement issued by TÜV Rheinland Immissionsschutz und Energiesysteme GmbH dated 31 March 2009

Publication in the German Federal Gazette: BAnz. 26. Januar 2011, Nr. 14, S. 294, Chap. IV notification 5, Announcement by UBA dated 10 January 2011:

5 Notification as regards Federal Environment Agency (UBA) notices of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.1) and of 3 August 2009 (BAnz. p. 2929, chapter III notification 1)

The APMA 370 measuring system for CO manufactured by Horiba Ltd, Japan, and Horiba Europe GmbH meets the requirements defined in standard EN 14626. Furthermore, the manufacturing process and the quality management for the APMA 370 measuring system meet the requirements of EN 15267. The performance test report is available online at www.qal1.de.

Statement of by TÜV Rheinland Energie und Umwelt GmbH dated 6 October 2010

Publication in the German Federal Gazette: BAnz. 02. März 2012, Nr. 36, S. 920, Chap. V notification 16, Announcement by UBA dated 23 February 2012:

16 Notification as regards Federal Environment Agency (UBA) notices of 21 February 2006 (p. 2653, chapter IV number 2.1) and of 10 January 2011 (BAnz. p. 294, chapter IV notification 5)

There is an addendum to test report No. 936/21204643/B for the APMA 370 measuring system for CO manufactured by Horiba, Ltd., Japan and Horiba Europe GmbH. The addendum is assigned report No. 936/21204643/B1 and after its publication is an integral part of the test report No. 936/21204643/B and is also available online at www.qal1.de.

Statement of by TÜV Rheinland Energie und Umwelt GmbH dated 3 November 2011

Publication in the German Federal Gazette: BAnz AT 05.03.2013 B10, Chap. V notification 7, Announcement by UBA dated 12 February 2013:

7 Notification as regards Federal Environment Agency (UBA) notices of 21 February 2006 (p. 2653, chapter IV number 2.1) and of 23 February 2012 (BAnz. p. 920, chapter V notification 16)

The APMA 370 measuring system for CO manufactured by Horiba Ltd, Japan, and Horiba Europe GmbH may optionally be equipped with an additional calibration port. Calibration gas may be fed upstream or downstream of the sample gas filter using a three-way valve.

Statement of by TÜV Rheinland Energie und Umwelt GmbH dated 11 October 2012

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, Chap. VI
notification 26, Announcement by UBA dated 27 February 2014:

**26 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (p. 2653, chapter IV number 2.1) and
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter V notification 7)**

The APMA 370 measuring system for CO manufactured by Horiba Ltd, Japan and Horiba Europe GmbH meets the requirements defined in standard EN 14626 (December 2012 version).

An addendum as integral part of test report No. 936/21222689/B is available online at www.qal1.de.

Statement of TÜV Rheinland Energie und Umwelt GmbH dated 5 October 2013

Publication in the German Federal Gazette: BAnz AT 01.08.2016 B11, Chap. V
notification 30, Announcement by UBA dated 14 July 2016:

**30 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (p. 2653, chapter IV number 2.1) and
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter VI notification 26)**

The APMA-370 measuring system for CO manufactured by HORIBA Ltd. is equipped with a new display which, in design and functionality, largely corresponds to its predecessor. In addition, the power supply ZWS-BAF may also be used.

The current software version of the measuring system is:
P1000878001K

Statement of TÜV Rheinland Energie und Umwelt GmbH dated 29 February 2016.

Publication in the German Federal Gazette: BAnz AT 22.07.2019 B8, Chap. V notification 9,
Announcement by UBA dated 28 June 2019:

**9 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.1) and
of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V notification 30)**

The latest software version of the APMA-370 measuring system for CO manufactured by HORIBA Ltd. is:
P1000878001L

The rear of the housing was modified to cater for additional cable connections.

Statement of TÜV Rheinland Energy GmbH dated 5 March 2019

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV
notification 53, Announcement by UBA dated 24 February 2020:

**53 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (p. 2653, chapter IV number 2.1) and
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter V 9th notification)**

The latest software version of the APMA-370 measuring system for CO
manufactured by HORIBA Ltd. is:
P1000878001M

Statement issued by TÜV Rheinland Energy GmbH dated 20 September 2019

Publication in the German Federal Gazette: BAnz AT 03.05.2021 B9, Chap. III
notification 35, Announcement by UBA dated 31 March 2021:

**35 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.1) and
of 24 February 2020 (BAnz AT 24.03.2020 B7, chapter IV notification 53)**

The latest software version of the APMA-370 measuring system for CO
manufactured by HORIBA Ltd. is:
P1000878001N.

Statement issued by TÜV Rheinland Energy GmbH dated 08 September 2020

Publication in the German Federal Gazette: BAnz AT 20.03.2023 B6, Chap. IV
notification 71, Announcement by UBA dated 21 February 2023:

**71 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 35)**

The current software version of the APMA-370 measuring system for CO
from the company HORIBA Ltd. is:
P1000878001P

The measuring system can also be equipped with an LCD display of type TL-0267L2
from the manufacturer AUO.

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2022

Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, Chap. IV
notification 77, Announcement by UBA dated 2 April 2025:

**77 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.1) and
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 71)**

The current software version of the APMA-370 measuring system for CO
from HORIBA Ltd. is:
P1000878001Q

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 26 September 2024

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The APMA 370 CO-Analyser operates based on the principle of non-dispersal infrared absorption.

The measuring principle complies with the reference measuring method described in section 5.2 of Standard EN 14626 (2012). The attenuation of infrared light is measured after passing through the test cuvette. According to Lambert-Beer law, attenuation is a measure of the concentration of CO within the cuvette.

The APMA-370 measuring system uses the modulation effect of infrared absorption within the test gas itself if zero and test gas are led to the measuring cell at specific flow rates. The switch is made via a solenoid valve which is clocked at a frequency of 1 Hz. As long as the concentration of the measured component within the cell does not change, the output of the detector is practically zero. Thus, no zero drift occurs.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet:
qal1.de.

History of documents

Certification of APMA 370 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/21204643/B dated 5 January 2006

TÜV Immissionsschutz und Energiesysteme GmbH

Publication: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV number 2.1

UBA announcement dated 21 February 2006

Notifications

Statement of TÜV Immissionsschutz und Energiesysteme GmbH dated 31 March 2009

Publication: BAnz. 25 August 2009, No. 125, p. 2929, chapter III notification 1

UBA announcement dated 3 August 2009

(Soft- and hardware changes)

Initial certification according to EN 15267

Certificate No. 0000028754_00: 9 February 2011

Expiry date of the certificate: 25 January 2016

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 6 October 2010

Test report: 936/21204643/B dated 5 January 2006

Publication: BAnz. 26 January 2011, No. 14, p. 294, chapter IV number 5

UBA announcement dated 10 January 2011

Certificate based on a notification

Certificate No. 0000028754_01: 16 March 2012

Expiry date of the certificate: 25 January 2016

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 3 November 2011

Test report: 936/21204643/B1 dated 27 July 2011

Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter V number 16

UBA announcement dated 23 February 2012

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 11 October 2012

Publication: BAnz AT 05.03.2013 B10, chapter V notification 7

UBA announcement dated 12 February 2013

(Additional hardware)

Certificate based on a notification

Certificate No. 0000028754_02: 29 April 2014

Expiry date of the certificate: 25 January 2016

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 5 October 2013

Test report: 936/21222689/B dated 5 October 2013

Publication: BAnz AT 01.04.2014 B12, chapter VI number 26

UBA announcement dated 27 February 2014

Renewal of certificates

Certificate No. 0000028754_03: 21 January 2016

Expiry date of the certificate: 25 January 2021

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 29 February 2016

Publication: BAuz AT 01.08.2016 B11, chapter V notification 30

UBA announcement dated 14 July 2016

(Soft- and hardware changes and new system name)

Statement issued by TÜV Rheinland Energy GmbH dated 5 March 2019

Publication: BAuz AT 22.07.2019 B8, chapter V notification 9

UBA announcement dated 28 June 2019

(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 20 September 2019

Publication: BAuz AT 24.03.2020 B7, chapter IV notification 53

UBA announcement dated 24 February 2020

(Software changes)

Renewal of certificates

Certificate No. 0000028754_04: 25 January 2021

Expiry date of the certificate: 25 January 2026

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 8 September 2020

Publication: BAuz AT 03.05.2021 B9, chapter III notification 35

UBA announcement dated 31 March 2021

(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2022

Publication: BAuz AT 20.03.2023 B6, chapter IV notification 71

UBA announcement dated 21 February 2023

(Soft- and hardware changes)

Statement of TÜV Rheinland Energy & Environment GmbH dated 26 September 2024

Publication: BAuz AT 19.05.2025 B3, chapter IV notification 77

UBA announcement dated 2 April 2025

(Software changes)

Renewal of certificates

Certificate No. 0000028754_05: 22 January 2026

Expiry date of the certificate: 25 January 2031

Expanded uncertainty from the results obtained in the laboratory tests for analyser 1

Measuring device:	Horiba APMA 370	Measured component:	CO	Serial No.:	SN 10031	8h-limit value:	8.62	μmol/mol
Performance characteristic								
No.		Performance criterion	Result	Partial uncertainty	Partial uncertainty	Partial uncertainty	Partial uncertainty	Partial uncertainty
1	Repeatability standard deviation at zero	≤ 0.3 μmol/mol	0.035	U _{r,z}	0.01	U _{r,z}	0.01	U _{r,z}
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 μmol/mol	0.064	U _r	0.00	U _r	0.00	U _r
3	"lack of fit" at 8hlimit value	≤ 4.0% of measured value	0.700	U _i	0.03	U _i	0.03	U _i
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 μmol/mol/KPa	0.006	U _{gp}	0.02	U _{gp}	0.02	U _{gp}
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 μmol/mol/K	0.005	U _{gt}	0.01	U _{gt}	0.01	U _{gt}
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 μmol/mol/K	0.077	U _{st}	0.20	U _{st}	0.20	U _{st}
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 μmol/mol/V	0.006	U _V	0.06	U _V	0.06	U _V
8a	Interferent H ₂ O with 21 nmol/mol	≤ 1.0 μmol/mol (Zero)	0.233	U _{H2O}	0.02	U _{H2O}	0.02	U _{H2O}
8b	Interferent CO ₂ with 500 μmol/mol	≤ 1.0 μmol/mol (Span)	-0.009	U _{int, pos}	0.017	U _{int, pos}	0.017	U _{int, pos}
8c	Interferent NO with 1 μmol/mol	≤ 0.5 μmol/mol (Zero)	-0.164	U _{int, neg}	-0.026	U _{int, neg}	-0.026	U _{int, neg}
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 μmol/mol (Span)	-0.026	U _{int, pos}	-0.353	U _{int, pos}	-0.353	U _{int, pos}
9	Averaging effect	≤ 7.0% of measured value	1.250	U _{av}	0.06	U _{av}	0.06	U _{av}
18	Difference sample/calibration port	≤ 1.0%	0.000	U _{usc}	0.00	U _{usc}	0.00	U _{usc}
21	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	0.09	U _{cg}	0.09	U _{cg}
Combined standard uncertainty								
				U _c	0.2582	U _c	0.2582	U _c
Expanded uncertainty								
				U	0.5165	U	0.5165	U
Relative expanded uncertainty								
				W	5.99	W	5.99	W
Maximum allowed expanded uncertainty								
				W _{req}	15	W _{req}	15	W _{req}

Expanded uncertainty from the results obtained in the laboratory tests for analyser 2

Measuring device:	Horiba APMA 370	Measured component:	CO	Serial-No.:	SN 10032	8h-limit value:	8.62	µmol/mol
Performance characteristic								
No.		Performance standard deviation at zero	≤	0.3 µmol/mol	0.028	Partial uncertainty		Square of partial uncertainty
1		Repeatability	≤	0.4 µmol/mol	0.070	u _{r,z}	0.01	0.0000
2		Repeatability standard deviation at 8h-limit value	≤	4.0% of measured value	0.800	u _r	0.00	0.0000
3		"lack of fit" at 8h-limit value	≤	0.7 µmol/mol/kPa	0.009	u _{gp}	0.02	0.0016
4		Sensitivity coefficient of sample gas pressure at 8h-limit value	≤	0.3 µmol/mol/K	0.004	u _{gt}	0.01	0.0005
5		Sensitivity coefficient of sample gas temperature at 8h-limit value	≤	0.3 µmol/mol/K	0.067	u _{st}	0.17	0.0304
6		Sensitivity coefficient of surrounding temperature at 8h-limit value	≤	0.3 µmol/mol/K	-0.007	u _v	-0.06	0.0041
7		Sensitivity coefficient of electrical voltage at 8h-limit value	≤	1.0 µmol/mol (Zero)	0.181	u _{H2O}	0.01	0.0001
8a		Interferent H ₂ O with 21 mmol/mol	≤	1.0 µmol/mol (Span)	-0.009			
8b		Interferent CO ₂ with 500 µmol/mol	≤	0.5 µmol/mol (Zero)	0.017	u _{int, pos}		
8c		Interferent NO with 1 µmol/mol	≤	0.5 µmol/mol (Span)	-0.086			
8d		Interferent N ₂ O with 50 nmol/mol	≤	0.5 µmol/mol (Zero)	-0.009	u _{int, neg}	0.01	0.0000
9		Averaging effect	≤	0.5 µmol/mol (Span)	0.017			
18		Difference sample/calibration port	≤	7.0% of measured value	-1.100	u _{av}	-0.05	0.0030
21		Uncertainty of test gas	≤	1.0%	0.000	u _{sc}	0.00	0.0000
			≤	3.0%	2.000	u _{cg}	0.09	0.0074
				Combined standard uncertainty	u _c			0.2176
				Expanded uncertainty	U			0.4353
				Relative expanded uncertainty	W			5.05
				Maximum allowed expanded uncertainty	W _{req}			15
								%

**Expanded uncertainty from the results obtained in the laboratory
and field tests for analyser 1**

Measuring device:	Honiba APMA 370	Measured component:	CO	Serial-No.:	SN 10031	8h-limit value:	8.62	μmol/mol
No.		Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 0.3 μmol/mol	0.035	u _{r,z}	0.01	0.0001		
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 μmol/mol	0.064	u _r	not considered, as u _r < u _{r,f}	-		
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.700	u _f	0.03	0.0012		
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 μmol/mol/kPa	0.006	u _{gp}	0.02	0.0002		
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 μmol/mol/K	0.005	u _{gt}	0.01	0.0002		
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 μmol/mol/K	0.077	u _{st}	0.20	0.0398		
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 μmol/mol/V	0.006	u _v	0.06	0.0030		
8a	Interferent H ₂ O with 21 μmol/mol	≤ 1.0 μmol/mol (Zero)	-0.009	u _{H2O}	0.02	0.0003		
8b	Interferent CO ₂ with 500 μmol/mol	≤ 1.0 μmol/mol (Span)	0.233					
8c	Interferent NO with 1 μmol/mol	≤ 0.5 μmol/mol (Zero)	0.017	u _{int, pos}				
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 μmol/mol (Span)	-0.164	u _{int, neg}				
9	Averaging effect	≤ 7.0% of measured value	1.250	u _{av}	0.06	0.0039		
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.420	u _{r,f}	0.29	0.0869		
11	Long term drift at zero level	≤ 0.5 μmol/mol	-0.172	u _{dl,r,z}	-0.10	0.0099		
12	Long term drift at span level	≤ 5.0% of max. of certification range	-1.750	u _{dl,r,8h}	-0.09	0.0076		
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{asc}	0.00	0.0000		
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{eg}	0.09	0.0074		
		Combined standard uncertainty		u _c		0.4136	μmol/mol	
		Expanded uncertainty		U		0.8271	μmol/mol	
		Relative expanded uncertainty		W		9.60	%	
		Maximum allowed expanded uncertainty		W _{req}		15	%	

**Expanded uncertainty from the results obtained in the laboratory
and field tests for analyser 2**

Measuring device:	HORIBA APMA 370	Serial-No.:	SN 10032
Measured component:	CO	8h-limit value:	8.62
No.	Performance characteristic	Performance criterion	Result
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.028
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.070
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.800
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.009
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.004
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.067
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	-0.007
8a	Interferent H ₂ O with 21 µmol/mol	≤ 1.0 µmol/mol (Zero)	-0.009
8b	Interferent CO ₂ with 500 µmol/mol	≤ 1.0 µmol/mol (Span)	0.181
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Zero)	0.017
8d	Interferent N ₂ O with 50 µmol/mol	≤ 0.5 µmol/mol (Zero)	-0.017
9	Averaging effect	≤ 7.0% of measured value	-1.100
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.420
11	Long term drift at zero level	≤ 0.5 µmol/mol	-0.164
12	Long term drift at span level	≤ 5.0% of max. of certification range	-1.610
18	Difference sample/calibration port	≤ 1.0%	0.000
21	Uncertainty of test gas	≤ 3.0%	2.000
Combined standard uncertainty		U _c	0.3869
Expanded uncertainty		U	0.7737
Relative expanded uncertainty		W	8.98
Maximum allowed expanded uncertainty		W _{req}	15 %