

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000028757\_05

**Certified AMS:** APSA 370 for SO<sub>2</sub>

**Manufacturer:** Horiba Ltd.  
2 Miyanoigashi  
Kisshoin Minami-ku / Kyoto 610-8510  
Japan

**Test Institute:** TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested  
and found to comply with the standards  
VDI 4202-1 (2010), VDI 4203-3 (2010), EN 14212 (2012),  
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate  
(this certificate contains 14 pages).

The present certificate replaces certificate 0000028757\_04 dated 25 January 2021.




Publication in the German Federal Gazette  
(BAnz) of 14 October 2006


This certificate will expire on:  
25 January 2031

German Environment Agency

TÜV Rheinland  
Energy & Environment GmbH  
Cologne, 21 January 2026

Dessau, 22 January 2026

  
Dr. Marcel Langner  
Head of Section II 4

  
ppa. Dr. Peter Wilbring

<a href="http://www.umwelt-tuv.eu">www.umwelt-tuv.eu</a> qal1-info@tuv.com Tel. + 49 221 806-5200	TÜV Rheinland Energy & Environment GmbH Am Grauen Stein 51105 Köln
Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.	

**Test report:** 936/21204643/D dated 7 July 2006  
**Initial certification:** 26 January 2011  
**Expiry date:** 25 January 2031  
**Certificate:** Renewal (of previous certificate 0000028757\_04 of  
25 January 2021 valid until 25 January 2026)  
**Publication:** BAnz. 14 October 2006, No. 194, p. 6715, chapter IV No. 2.1

### **Approved application**

The tested AMS is suitable for continuous immission measurement of SO<sub>2</sub> in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a four month field test.

The AMS is approved for an ambient temperature range of 0 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

### **Basis of the certification**

This certification is based on:

- Test report 936/21204643/D dated 7 July 2006 of TÜV Immissionsschutz und Energiesysteme GmbH
- Addenda 936/21204643/D1 dated 27 July 2011 and 936/21222689/D dated 5 October 2013
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 14 October 2006, No. 194, p. 6715, chapter IV No. 2.1, Announcement by UBA dated 12 September 2006:

**AMS designation:**

APSA 370

**Manufacturer:**

HORIBA, Ltd., Kyoto, Japan

Distribution: HORIBA Europe GmbH, Leichlingen

**Field of application:**

For continuous ambient air monitoring of sulphur dioxide (stationary operation)

**Measuring ranges during the performance test:**

Component	Measurement ranges	Unit
SO <sub>2</sub>	0 - 700	µg/m <sup>3</sup>
	0 - 1,000	µg/m <sup>3</sup>

**Software version:**

P1000878001C

**Test institute:**

TÜV Immissionsschutz und Energiesysteme GmbH, Cologne

TÜV Rheinland Group

**Report No.:**

936/21204643/D dated 7 July 2006

Publication in the German Federal Gazette: BAnz. 25. August 2009, Nr. 125, S. 2929, Chap. III notification 4, Announcement by UBA dated 3 August 2009:

**4 Notification as regards Federal Environment Agency notice of 12 September 2006 (BAnz. p. 6717)**

The latest software version of the APSA 370 ambient air measuring system manufactured by Horiba Europe GmbH is:  
P1000878001J

The type GD-6 EH sample gas pump manufactured by Horiba may be used instead of the N 86.0 KNE sample gas pump manufactured by KNF.

Statement issued by TÜV Immissionsschutz und Energiesysteme GmbH dated 31 March 2009

Publication in the German Federal Gazette: BAnz. 02. März 2012, Nr. 36, S. 920, Chap. IV notification 7, Announcement by UBA dated 23 February 2012:

**7 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and of 3 August 2009 (BAnz. p. 2929, chapter III notification 4)**

The APSA 370 measuring system for SO<sub>2</sub> manufactured by Horiba Ltd, Japan, and Horiba Europe GmbH meets the requirements defined in standard EN 14212. Furthermore, the manufacturing process and the quality management for the Model APSA 370 for SO<sub>2</sub> measuring system meet the requirements of EN 15267. The test report on performance testing is available on the internet at [www.qal1.de](http://www.qal1.de).

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 6-10-2010

Publication in the German Federal Gazette: BAnz. 02. März 2012, Nr. 36, S. 920, Chap. V notification 18, Announcement by UBA dated 23 February 2012:

**18 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (p. 6715, chapter IV number 2.1) and of 10 January 2011 (BAnz. p. 294, chapter IV notification 7)**

There is an addendum to test report No. 936/21204643/D for the APSA 370 measuring system for SO<sub>2</sub> manufactured by Horiba, Ltd., Japan and Horiba Europe GmbH. The addendum is assigned report No. 936/21204643/D1 and after its publication is an integral part of the test report No. 936/21204643/D and is also available online at [www.qal1.de](http://www.qal1.de).

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 3-11-2011

Publication in the German Federal Gazette: BAnz AT 05.03.2013 B10, Chap. V notification 10, Announcement by UBA dated 12 February 2013:

**10 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (p. 6715, chapter IV number 2.1) and of 23 February 2012 (BAnz. p. 920, chapter V notification 18)**

The APSA 370 measuring system for SO<sub>2</sub> manufactured by Horiba Ltd, Japan, and Horiba Europe GmbH may optionally be equipped with an additional calibration port. Calibration gas may be fed upstream or downstream of the sample gas filter using a three-way valve.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH  
dated 11 October 2012

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, Chap. VI  
notification 29, Announcement by UBA dated 27 February 2014:

**29 Notification as regards Federal Environment Agency (UBA) notices  
of 12 September 2006 (p. 6715, chapter IV number 2.1) and  
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter V notification 10)**

The APSA 370 measuring system for SO<sub>2</sub> manufactured by Horiba Ltd, Japan, and Horiba Europe GmbH meets the requirements defined in standard EN 14212 (November 2012 version). An addendum as integral part of test report No. 936/21222689/D is available online at [www.qal1.de](http://www.qal1.de). In addition to the reference diode type S7798 used so far for measuring the UV light intensity a type S12698 (TO5) diode may be used.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH  
dated 8 October 2013

Publication in the German Federal Gazette: BAnz AT 01.08.2016 B11, Chap. V  
notification 33, Announcement by UBA dated 14 July 2016:

**33 Notification as regards Federal Environment Agency (UBA) notices  
of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter VI notification 29)**

The APSA-370 measuring system for SO<sub>2</sub> manufactured by HORIBA Ltd. is equipped with a new display which, in design and functionality, largely corresponds to its predecessor. In addition, the power supply ZWS-BAF may also be used.

The current software version of the APSA-370 measuring system for SO<sub>2</sub> is:  
P1000878001K

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 29-2-2016.

Publication in the German Federal Gazette: BAnz AT 22.07.2019 B8, Chap. V notification 12, Announcement by UBA dated 28 June 2019:

**12 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V notification 33)**

The latest software version of the APSA-370 for SO<sub>2</sub> manufactured by HORIBA Ltd. is:  
P1000878001L

The rear of the housing was modified to cater for additional cable connections. Instead of the xenon lamp type L4646 with opaque socket, the new version of this lamp with a transparent socket and the same designation may be used.

Statement issued by TÜV Rheinland Energy GmbH dated 5 March 2019

Publication in the German Federal Gazette: BAnz AT 07.05.2020 B8, Chap. III notification 3, Announcement by UBA dated 31 March 2020:

**3 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (p. 6715, chapter IV number 2.1) and of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter V notification 12)**

The latest software version of the APSA-370 for SO<sub>2</sub> manufactured by HORIBA Ltd. is:  
P1000878001M

The xenon lamp may optionally be equipped with an additional fixation.

Statement issued by TÜV Rheinland Energy GmbH dated 06 December 2019

Publication in the German Federal Gazette: BAnz AT 03.05.2021 B9, Chap. III notification 38, Announcement by UBA dated 31 March 2021:

**38 Notification as regards Federal Environment Agency (UBA) notices of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and of 31 March 2020 (BAnz AT 07.05.2020 B8, chapter III notification 3)**

The latest software version of the APSA-370 measuring system for SO<sub>2</sub> manufactured by HORIBA Ltd. is:  
P1000878001N.

Statement issued by TÜV Rheinland Energy GmbH dated 08 September 2020

Publication in the German Federal Gazette: BAnz AT 11.04.2022 B10, Chap. VI  
notification 15, Announcement by UBA dated 9 March 2022:

**18 Notification as regards Federal Environment Agency (UBA) notices  
of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and  
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 38)**

The immission measuring device APSA-370 for SO<sub>2</sub> of the company HORIBA Ltd. can be equipped with a new heater for the measuring cell block in the future. The new heater has the part number 3200837467.

Statement issued by TÜV Rheinland Energy GmbH dated 12 September 2021

Publication in the German Federal Gazette: BAnz AT 20.03.2023 B6, Chap. IV  
notification 74, Announcement by UBA dated 21 February 2023:

**74 Notification as regards Federal Environment Agency (UBA) notices  
of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and  
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 15)**

The current software version of the APSA-370 measuring system for SO<sub>2</sub> from the company HORIBA Ltd. is:  
P1000878001P

The measuring system can also be equipped with an LCD display of the type TL-0267L2 from the manufacturer AUO.

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2022

Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, Chap. IV  
notification 80, Announcement by UBA dated 2 April 2025:

**80 Notification as regards Federal Environment Agency (UBA) notices  
of 12 September 2006 (BAnz. p. 6715, chapter IV number 2.1) and  
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 74)**

The current software version of the APSA-370 measuring system for SO<sub>2</sub> from HORIBA Ltd. is  
P1000878001Q

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 26-9-2024

**Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The APSA 370 SO<sub>2</sub> analyser measures sulphur dioxide by means of ultraviolet fluorescence.

The measuring principle corresponds to the setup and functioning of the measuring principle described in EN 14212 (2012) section 5.2 for the component SO<sub>2</sub>.

The sample air first passes a filter in the APSA 370, where coarse dirt particles are filtered out. The sample is drawn through the appropriate inlet. First, the sampled air is drawn through a hydrocarbon scrubber to remove any interference by aromatic hydrocarbons that may be present. SO<sub>2</sub> molecules remain unaffected. The sample is then introduced into a reaction chamber, where it is irradiated by UV light in the wave-length range between 200 nm and 220 nm. The radiation excites the SO<sub>2</sub> molecules energetically. The 4 mirrors upstream of UV lamp only allow those wavelengths to pass which excite the SO<sub>2</sub> molecules.

When the SO<sub>2</sub> molecules return to their lower energy state, they emit UV fluorescent light in a wavelength range of 240 nm to 420 nm. Only this light reaches the photomultiplier (PMT) after having passed the filter. The PMT measures the UV emission and converts it to an electric signal. A photo detector at the end of the reaction chamber continuously measures the lamp radiation and corrects the measuring result in the event of fluctuations.

**General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet:  
**qal1.de**

### **History of documents**

Certification of APSA 370 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

#### **Basic test**

Test report: 936/21204643/D dated 7 July 2006  
TÜV Immissionsschutz und Energiesysteme GmbH  
Publication: BAnz. 14 October 2006, No. 194, p. 6715, chapter IV number 2.1  
UBA announcement dated 12 September 2006

#### **Notifications**

Statement issued by TÜV Immissionsschutz und Energiesysteme GmbH  
dated 31 March 2009  
Publication: BAnz. 25 August 2009, No. 125, p. 2929, chapter III notification 4  
UBA announcement dated 3 August 2009  
(Soft- and hardware changes)

#### **Initial certification according to EN 15267**

Certificate No. 0000028757\_00: 9 February 2011  
Expiry date of the certificate: 25 January 2016  
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 6 October 2010  
Test report: 936/21204643/D dated 7 July 2006  
Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter IV number 7  
UBA announcement dated 23 February 2012

#### **Certificate based on a notification**

Certificate No. 0000028757\_01: 16 March 2012  
Expiry date of the certificate: 25 January 2016  
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 3 November 2011  
Test report: 936/21204643/D1 dated 27 July 2011  
Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter V number 18  
UBA announcement dated 23 February 2012

#### **Notifications**

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 11 October 2012  
Publication: BAnz AT 05.03.2013 B10, chapter V notification 10  
UBA announcement dated 12 February 2013  
(Hardware change)

#### **Certificate based on a notification**

Certificate No. 0000028757\_02: 29 April 2014  
Expiry date of the certificate: 25 January 2016  
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 8 October 2013  
Test report: 936/21222689/D dated 5 October 2013  
Publication: BAnz AT 01.04.2014 B12, chapter VI number 29  
UBA announcement dated 27 February 2014

#### **Renewal of certificates**

Certificate No. 0000028757\_03: 21 January 2016  
Expiry date of the certificate: 25 January 2021

### **Notifications**

Statement issued by TÜV Rheinland Energy GmbH dated 29 February 2016

Publication: BAuz AT 01.08.2016 B11, chapter V notification 33

UBA announcement dated 14 July 2016

(Soft- and hardware changes and new system name)

Statement issued by TÜV Rheinland Energy GmbH dated 5 March 2019

Publication: BAuz AT 22.07.2019 B8, chapter V notification 12

UBA announcement dated 28 June 2019

(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 6 December 2019

Publication: BAuz AT 07.05.2020 B8, chapter III notification 3

UBA announcement dated 31 March 2020

(Soft- and hardware changes)

### **Renewal of certificates**

Certificate No. 0000028757\_04: 25 January 2021

Expiry date of the certificate: 25 January 2026

### **Notifications**

Statement issued by TÜV Rheinland Energy GmbH dated 8 September 2020

Publication: BAuz AT 03.05.2021 B9, chapter III notification 38

UBA announcement dated 31 March 2021

(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 12 September 2021

Publication: BAuz AT 11.04.2022 B10, chapter VI notification 15

UBA announcement dated 9 March 2022

(Hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2022

Publication: BAuz AT 20.03.2023 B6, chapter IV notification 74

UBA announcement dated 21 February 2023

(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy & Environment GmbH

dated 26 September 2024

Publication: BAuz AT 19.05.2025 B3, chapter IV notification 80

UBA announcement dated 2 April 2025

(Software changes)

### **Renewal of certificates**

Certificate No. 0000028757\_05: 22 January 2026

Expiry date of the certificate: 25 January 2031

## Expanded uncertainty laboratory, system 1

Measuring device:	Horiba APSA 370		
Measured component:	SO <sub>2</sub>		
Serial-No.:	SN 10012		
No.	Performance characteristic	Performance criterion	Result
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.070
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.140
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.600
4	Sensitivity coefficient of sample gas, pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.020
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	-0.010
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.060
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	-0.010
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.070
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	-1.500
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.100
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	0.030
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Span)	-0.500
9	Averaging effect	≤ 7.0% of measured value	-3.600
18	Difference sample/calibration port	≤ 1.0%	0.000
21	Uncertainty of test gas	≤ 3.0%	2.000
Combined standard uncertainty			U <sub>c</sub>
Expanded uncertainty			U
Relative expanded uncertainty			W
Maximum allowed expanded uncertainty			W <sub>max</sub>
			mmol/mol
1h-limit value:			132
			nmol/mol

**Expanded uncertainty laboratory, system 2**

Measuring device:	Horiiba APSA 370	Measured component:	SO2	Serial-No.:	SN 10011	1h-limit value:	132	nmol/mol
<b>Performance characteristic</b>								
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty			
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.080	$u_{1,2}$	0.02			0.0006
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.130	$u_{1,h}$	0.36			0.1315
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	-0.800	$u_{1,h}$	-0.61			0.3717
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.010	$u_{\text{up}}$	0.09			0.0073
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	-0.090	$u_{\text{gt}}$	-0.80			0.6361
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.250	$u_{\text{st}}$	2.22			4.9081
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.020	$u_{\text{v}}$	0.20			0.0405
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.070	$u_{\text{H2O}}$	-0.59			0.3432
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	-1.530					
		≤ 5.0 nmol/mol (Zero)	-0.030	$u_{\text{int, pos}}$				
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	0.570					
		≤ 5.0 nmol/mol (Zero)	0.000					
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	-0.270					
		≤ 5.0 nmol/mol (Zero)	3.100					
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.100	or	2.22			4.9344
		≤ 5.0 nmol/mol (Span)	-0.930					
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Zero)	0.030					
9	Averaging effect	≤ 10 nmol/mol (Span)	0.230	$u_{\text{int, neg}}$				
18	Difference sample/calibration port	≤ 7.0% of measured value	-4.300	$u_{\text{av}}$	-3.28			10.7390
21	Uncertainty of test gas	≤ 3.0%	2.000	$u_{\text{sc}}$	0.00			0.0000
<b>Combined standard uncertainty</b>								
				$u_c$	1.32			4.8841
				$U$				9.7683
				$W$				7.40
				$W_{\text{eq}}$	15			%

**Combined uncertainty, laboratory and field, system 1**

Measuring device:	Horiba APSA 370	Measured component:	SO <sub>2</sub>	Serial-No.:	SN 10012
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.070	u <sub>r,z</sub>	0.02
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.140	u <sub>r,lin</sub>	not considered, as u <sub>r,lin</sub> = 0.38 < u <sub>r,f</sub>
3	"lack of fit" at 1h-limit value	4.0% of measured value	0.600	u <sub>l,lin</sub>	0.46
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	2.0 nmol/mol/kPa	0.020	u <sub>gp</sub>	0.17
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	1.0 nmol/mol/K	-0.010	u <sub>gt</sub>	-0.09
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	1.0 nmol/mol/K	0.060	u <sub>st</sub>	0.54
7	Sensitivity coefficient of electrical voltage at 1h-limit value	0.30 nmol/mol/V	-0.010	u <sub>v</sub>	-0.10
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	10 nmol/mol (Zero)	0.070		
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	10 nmol/mol (Span)	-1.500	u <sub>H2O</sub>	-0.56
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	5.0 nmol/mol (Zero)	-0.100		
8d	Interferent NO with 500 nmol/mol	5.0 nmol/mol (Span)	0.030	u <sub>int, pos</sub>	0.030
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	5.0 nmol/mol (Zero)	0.000		
8f	Interferent m-Xylene with 1 µmol/mol	10 nmol/mol (Span)	-0.400	u <sub>int, neg</sub>	3.010
9	Averaging effect	7.0% of measured value	-3.600	u <sub>av</sub>	-2.74
10	Reproducibility standard deviation under field conditions	5.0% of average over 3 months	4.240	u <sub>r,f</sub>	5.60
11	Long term drift at zero level	4.0 nmol/mol	0.800	u <sub>d,l,z</sub>	0.46
12	Long term drift at span level	5.0% of max. of certification range	1.310	u <sub>d,lin</sub>	1.00
13	Difference sample/calibration port	1.0%	0.000	u <sub>asc</sub>	0.000
21	Uncertainty of test gas	3.0%	2.000	u <sub>cg</sub>	1.32
				Combined standard uncertainty	u <sub>c</sub>
				Expanded uncertainty	U
				Relative expanded uncertainty	W
				Maximum allowed expanded uncertainty	W <sub>req</sub>

**Combined uncertainty, laboratory and field, system 2**

No.	Performance characteristic		Performance criterion		Result	Partial uncertainty	Square of partial uncertainty	1h-limit value: nmol/mol	Serial-No.: SN 10011 132
	Repeatability	standard deviation at 1h-limit value	≤	1.0 nmol/mol					
1	Repeatability standard deviation at zero	≤	3.0 nmol/mol	0.130	$u_{r,h}$	$u_{r,z}$	0.02	0.0006	
2	Repeatability standard deviation at 1h-limit value	≤	3.0 nmol/mol	0.080	$u_{r,h}$	$u_{r,z}$	0.02	0.0006	
3	"lack of fit" at 1h-limit value	≤	4.0% of measured value	-0.800	$u_{l,h}$	-0.61	-	0.3717	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤	2.0 nmol/mol/kPa	0.010	$u_{g,p}$	0.09	0.0073		
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤	1.0 nmol/mol/K	-0.090	$u_{g,t}$	-0.80	0.6361		
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤	1.0 nmol/mol/K	0.250	$u_{s,t}$	2.22	4.9081		
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤	0.30 nmol/mol/V	0.020	$u_V$	0.20	0.0405		
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤	10 nmol/mol (Zero)	0.070	$u_{H_2O}$	-1.530	-		
8b	Interferent H <sub>2</sub> S with 21 nmol/mol	≤	10 nmol/mol (Span)	-0.59	$u_{H_2O}$	-0.59	0.3432		
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤	5.0 nmol/mol (Zero)	-0.030	$u_{NH_3, pos}$	0.570	0.3132		
8d	Interferent NO with 500 nmol/mol	≤	5.0 nmol/mol (Zero)	0.000	$u_{NO}$	-0.270	-		
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤	5.0 nmol/mol (Zero)	3.100	$u_{NO_2}$	3.710	4.9344		
8f	Interferent m-Xylene with 1 µmol/mol	≤	10 nmol/mol (Zero)	-0.930	$u_{m-Xylene}$	2.22	-		
9	Averaging effect	≤	7.0% of measured value	0.100	$u_{avg}$	-3.28	10.7390		
10	Reproducibility standard deviation under field conditions	≤	5.0% of average over 3 months	4.240	$u_{r,f}$	5.60	31.3242		
11	Long term drift at zero level	≤	4.0 nmol/mol	0.700	$u_{d,l,z}$	0.40	0.1633		
12	Long term drift at span level	≤	5.0% of max. of certification range	0.630	$u_{d,l,h}$	0.48	0.2305		
18	Difference sample/calibration port	≤	1.0%	0.000	$u_{usc}$	0.00	0.0000		
21	Uncertainty of test gas	≤	3.0%	2.000	$u_{cg}$	1.32	1.7424		
			Combined standard uncertainty	$u_c$			7.4459	nmol/mol	
			Expanded uncertainty	U			14.8918	nmol/mol	
			Relative expanded uncertainty	W			11.28	%	
			Maximum allowed expanded uncertainty	W <sub>req</sub>			15	%	