

CONFIRMATION

Notification: 0000043107_03_01
on changes according to EN 15267 regarding certificate 0000043107_03 dated 20 March 2024

AMS designation: APDA-372 / APDA-372 E for PM₁₀ and PM_{2,5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**54 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and
of 5 July 2023 (BAnz AT 02.08.2023 B7, chapter III notification 36)**


The current software version for the APDA-372 and APDA-372 E measuring systems
for PM₁₀ and PM_{2,5} from HORIBA Europe GmbH is:
100539.0014.0001.0001.0011


Firmware 400045_FW_V0314 can also be used for the compact weather station from
Thies.

The circuit board layout of the MMS-PT100 has changed to Rev. 7.1

Statement issued by TÜV Rheinland Energy GmbH dated 01 September 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.