

# CONFIRMATION

Notification: 0000037055\_03\_01  
on changes according to EN 15267 regarding certificate 0000037055\_03 dated 20 March 2024

---

**AMS designation:** BAM 1020 PM<sub>10</sub> mit PM<sub>10</sub>-Vorabscheider for PM<sub>10</sub>

**Manufacturer:** Met One Instruments Inc.  
1600, Washington Blvd.  
Grants Pass, Oregon 97526,  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 19 March 2024  
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**55 Notification as regards Federal Environment Agency (UBA) notices  
of 12 April 2007 (BAnz. p. 4139, chapter III number 1.2) and  
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 78)**

The current software versions for the BAM-1020 measuring system for PM<sub>10</sub> particulate  
matter from Met One Instruments, Inc. are as follows:

Version without touch screen 3236-05 3.14.4.

Version with touch-sensitive screen (old screen version) 3236-77 V5.2.0.


Version with touch-sensitive screen (new screen version) R9.2.2.


Software versions above R9.2.2 are obsolete.

Statement issued by TÜV Rheinland Energy GmbH dated 01 September 2023

---

TÜV Rheinland Energy & Environment GmbH  
Cologne, 24 May 2024

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.