

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000039319_02

Certified AMS: MGS300 for CO, SO₂, NO, NO₂, HCl, HF, CH₄, CO₂, H₂O, N₂O and NH₃

Manufacturer: MKS Instruments Inc.
651 Lowell Street,
Methuen, MA 01844
USA

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007
and EN 14181: 2004**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).

The present certificate replaces Certificate No. 0000039319_01 of 29 April 2014




Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000039319


Publication in the German Federal Gazette
(BAnz.) of 5 August 2014

German Federal Environment Agency
Dessau, 9 September 2014


i. A. Dr. Marcel Langner

This certificate will expire on:
22 July 2018

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 8 September 2014


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

| | |
|-------------------------------|---|
| Test report: | 936/21208291/C of 20 March 2014 |
| Initial certification: | 23 July 2013 |
| Expiry date: | 22 July 2018 |
| Publication: | BAnz AT 5 August 2014 B11, chapter I, no. 4.3 |

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III, at waste incineration plants according to Directive 2010/75/EU, chapter IV and other plants requiring official approval. The measured ranges have been selected considering the wide application range of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a twelve-month field test at a municipal waste incinerator.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21208291/C of 20 March 2014 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette: BAnz AT 5 August 2014 B11, chapter I, no. 4.3
UBA announcement of 17 July 2014

AMS designation:

MGS300 for CO, SO₂, NO, NO₂, HCl, HF, CH₄, CO₂, H₂O, N₂O and NH₃

Manufacturer:

MKS Instruments Inc., Methuen, USA

Field of application:

For measurements at plants requiring official approval (e.g. Directive 2010/75/EU on industrial emissions, chapters III and IV)

Measuring ranges during the performance test:

| Component | Certification range | Supplementary range | | Unit |
|------------------|---------------------|---------------------|----------|-------------------|
| HF | 0 - 3 | 0 - 10 | - | mg/m ³ |
| N ₂ O | 0 - 50 | 0 - 100 | 0 - 500 | mg/m ³ |
| CO | 0 - 75 | 0 - 300 | 0 - 1500 | mg/m ³ |
| SO ₂ | 0 - 75 | 0 - 300 | 0 - 2000 | mg/m ³ |
| NO | 0 - 200 | 0 - 400 | 0 - 1500 | mg/m ³ |
| NO ₂ | 0 - 50 | 0 - 100 | 0 - 1000 | mg/m ³ |
| HCl | 0 - 15 | 0 - 90 | 0 - 200 | mg/m ³ |
| NH ₃ | 0 - 10 | 0 - 75 | - | mg/m ³ |
| CO ₂ | 0 - 25 | - | - | Vol.-% |
| H ₂ O | 0 - 40 | - | - | Vol.-% |
| CH ₄ | 0 - 15 | 0 - 50 | 0 - 500 | mg/m ³ |

Software versions:

MG2000: 7.2

MGS300 Control: 01.04

Restriction:

The requirement of Standard EN 15267-3 for protection provided by enclosures was not met during performance testing. The measuring system shall be installed protected from dust and precipitation.

Notes:

1. The maintenance interval is six months.
2. Supplementary testing (extension of the maintenance interval) to the announcement of the Federal Environmental Agency (UBA) of 27 February 2014 (Federal Gazette (BAnz) AT 01.04.2014 B12, chapter I, no. 3.3).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21208291/C of 20 March 2014

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The MGS300 is a multi-component gas analysing system for continuous monitoring of exhaust gases at industrial incineration plants. The gas to be measured is extracted with help of a sample gas probe from the stack. Then the gas is forwarded with a heated sample line to the heated analyser system.

For the spectral acquisition of the gas concentration a Fourier-transformed infrared spectrometer is used. The measurement device consists of the following main components:

- FTIR analyser MKS type MultiGas 2030D-29805
- System cabinet with control computer, control electronics, gas supply and data output modules
- heated sample probe type JES301HFTIR
- heated sample gas line with stainless steel tubing, length during the type approval 10 meters
- heated sample gas pump type JHSS
- MGS300 Control software (for the control of general analyser functions, valve- and temperature control, visualisation of measured values)
- MG2000 software (interferometer control and calculation of measured values)

Automatic background measurement

The analysers performs a daily automatic zero adjustment with nitrogen. This adjustment lasts about 10 minutes.

Consumable gases

During the field test the measurement device was operated with nitrogen for the background cycle, with compressed air for the ejector-pump and with conditioned compressed air (dew point app. -40 °C and hydrocarbon free) for the interferometer purge.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of MGS300 for CO, SO₂, NO, NO₂, HCl, HF, CH₄, CO₂, H₂O, N₂O and NH₃ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate no. 0000039319: 20 August 2013
Expiry date of the certificate: 22 July 2018
Test report: 936/21208291/A of 26 March 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 23 July 2013 B4, chapter I, no. 3.2
UBA announcement of 3 July 2013

Supplementary testing according to EN 15267

Certificate no. 0000039319_01: 29 April 2014
Expiry date of the certificate: 22 July 2018
Test report: 936/21208291/B of 3 September 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 1 April 2014 B12, chapter I, no. 3.3
UBA announcement of 27 February 2014

Supplementary testing according to EN 15267

Certificate no. 0000039319_02: 9 September 2014
Expiry date of the certificate: 22 July 2018
Test report: 936/21208291/C of 20 March 2014
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 5 August 2014 B11, chapter I, no. 4.3
UBA announcement of 17 July 2014

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----------------|--------------------------|
| Certification range | CH ₄ | 0 - 15 mg/m ³ |
|---------------------|-----------------|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.27 mg/m ³ |
| Sum of negative CS at zero point | -0.12 mg/m ³ |
| Sum of positive CS at span point | 0.41 mg/m ³ |
| Sum of negative CS at span point | -0.42 mg/m ³ |
| Maximum sum of cross-sensitivities | -0.42 mg/m ³ |
| Uncertainty of cross-sensitivity | -0.242 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u ² |
|---|------------------|--------------------------|---|
| Repeatability standard deviation at set point * | u _r | 0.103 mg/m ³ | 0.011 (mg/m ³) ² |
| Lack of fit | u _{lof} | -0.058 mg/m ³ | 0.003 (mg/m ³) ² |
| Zero drift from field test | u _{d,z} | -0.078 mg/m ³ | 0.006 (mg/m ³) ² |
| Span drift from field test | u _{d,s} | 0.113 mg/m ³ | 0.013 (mg/m ³) ² |
| Influence of ambient temperature at span | u _t | 0.173 mg/m ³ | 0.030 (mg/m ³) ² |
| Influence of supply voltage | u _v | 0.074 mg/m ³ | 0.005 (mg/m ³) ² |
| Cross-sensitivity (interference) | u _i | -0.242 mg/m ³ | 0.059 (mg/m ³) ² |
| Influence of sample gas flow | u _p | -0.054 mg/m ³ | 0.003 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u _{rm} | 0.121 mg/m ³ | 0.015 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

| | | |
|---|------------------------------------|------------------------|
| Combined standard uncertainty (u _c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 0.38 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 0.75 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|--|---------|
| Requirement of 2010/75/EU | U in % of the ELV 10 mg/m ³ | 7.5 |
| Requirement of EN 15267-3 | U in % of the ELV 10 mg/m ³ | 30.0 ** |
| | U in % of the ELV 10 mg/m ³ | 22.5 |

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|----|--------------------------|
| Certification range | CO | 0 - 75 mg/m ³ |
|---------------------|----|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.00 mg/m ³ |
| Sum of negative CS at zero point | -2.12 mg/m ³ |
| Sum of positive CS at span point | 1.50 mg/m ³ |
| Sum of negative CS at span point | -1.30 mg/m ³ |
| Maximum sum of cross-sensitivities | -2.12 mg/m ³ |
| Uncertainty of cross-sensitivity | -1.225 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u^2 |
|--|-----------|--------------------------|---|
| Standard deviation from paired measurements under field conditions * | u_D | 0.245 mg/m ³ | 0.060 (mg/m ³) ² |
| Lack of fit | u_{lof} | 0.312 mg/m ³ | 0.097 (mg/m ³) ² |
| Zero drift from field test | $u_{d,z}$ | 0.260 mg/m ³ | 0.068 (mg/m ³) ² |
| Span drift from field test | $u_{d,s}$ | 0.346 mg/m ³ | 0.120 (mg/m ³) ² |
| Influence of ambient temperature at span | u_t | 0.379 mg/m ³ | 0.144 (mg/m ³) ² |
| Influence of supply voltage | u_v | 0.232 mg/m ³ | 0.054 (mg/m ³) ² |
| Cross-sensitivity (interference) | u_i | -1.225 mg/m ³ | 1.502 (mg/m ³) ² |
| Influence of sample gas flow | u_p | 0.271 mg/m ³ | 0.073 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 0.606 mg/m ³ | 0.368 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

| | | |
|---|------------------------------------|------------------------|
| Combined standard uncertainty (u_c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 1.58 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 3.09 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|--|-------------|
| Requirement of 2010/75/EU | U in % of the ELV 50 mg/m³ | 6.2 |
| Requirement of EN 15267-3 | U in % of the ELV 50 mg/m³ | 10.0 |
| | U in % of the ELV 50 mg/m³ | 7.5 |

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----------------|---------------|
| Certification range | CO ₂ | 0 - 25 Vol.-% |
|---------------------|-----------------|---------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | | |
|------------------------------------|-------|--------|
| Sum of positive CS at zero point | 0.00 | Vol.-% |
| Sum of negative CS at zero point | 0.00 | Vol.-% |
| Sum of positive CS at span point | 0.40 | Vol.-% |
| Sum of negative CS at span point | -0.30 | Vol.-% |
| Maximum sum of cross-sensitivities | 0.40 | Vol.-% |
| Uncertainty of cross-sensitivity | 0.231 | Vol.-% |

Calculation of the combined standard uncertainty

Tested parameter

| | | | | u^2 |
|--|-----------|--------|--------|-----------------------------|
| Standard deviation from paired measurements under field conditions * | u_D | 0.033 | Vol.-% | 0.001 (Vol.-%) ² |
| Lack of fit | u_{lof} | 0.058 | Vol.-% | 0.003 (Vol.-%) ² |
| Zero drift from field test | $u_{d,z}$ | 0.014 | Vol.-% | 0.000 (Vol.-%) ² |
| Span drift from field test | $u_{d,s}$ | 0.159 | Vol.-% | 0.025 (Vol.-%) ² |
| Influence of ambient temperature at span | u_t | 0.173 | Vol.-% | 0.030 (Vol.-%) ² |
| Influence of supply voltage | u_v | 0.118 | Vol.-% | 0.014 (Vol.-%) ² |
| Cross-sensitivity (interference) | u_i | 0.231 | Vol.-% | 0.053 (Vol.-%) ² |
| Influence of sample gas flow | u_p | -0.105 | Vol.-% | 0.011 (Vol.-%) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 0.202 | Vol.-% | 0.041 (Vol.-%) ² |

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

| | | | |
|---|------------------------------------|------|--------|
| Combined standard uncertainty (u_c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 0.42 | Vol.-% |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 0.83 | Vol.-% |

Relative total expanded uncertainty

| | | |
|---------------------------|------------------------------------|----------------|
| Requirement of 2010/75/EU | U in % of the ELV 25 Vol.-% | 3.3 |
| Requirement of EN 15267-3 | U in % of the ELV 25 Vol.-% | 10.0 ** |
| | U in % of the ELV 25 Vol.-% | 7.5 |

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|------------------|---------------|
| Certification range | H ₂ O | 0 - 40 Vol.-% |
|---------------------|------------------|---------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | | |
|------------------------------------|-------|--------|
| Sum of positive CS at zero point | 0.00 | Vol.-% |
| Sum of negative CS at zero point | 0.00 | Vol.-% |
| Sum of positive CS at span point | 0.70 | Vol.-% |
| Sum of negative CS at span point | -0.50 | Vol.-% |
| Maximum sum of cross-sensitivities | 0.70 | Vol.-% |
| Uncertainty of cross-sensitivity | 0.404 | Vol.-% |

Calculation of the combined standard uncertainty

Tested parameter

| | | | | u^2 |
|--|-----------|--------|--------|-----------------------------|
| Standard deviation from paired measurements under field conditions * | u_D | 0.127 | Vol.-% | 0.016 (Vol.-%) ² |
| Lack of fit | u_{lof} | 0.058 | Vol.-% | 0.003 (Vol.-%) ² |
| Zero drift from field test | $u_{d,z}$ | -0.046 | Vol.-% | 0.002 (Vol.-%) ² |
| Span drift from field test | $u_{d,s}$ | 0.300 | Vol.-% | 0.090 (Vol.-%) ² |
| Influence of ambient temperature at span | u_t | 0.265 | Vol.-% | 0.070 (Vol.-%) ² |
| Influence of supply voltage | u_v | 0.127 | Vol.-% | 0.016 (Vol.-%) ² |
| Cross-sensitivity (interference) | u_i | 0.404 | Vol.-% | 0.163 (Vol.-%) ² |
| Influence of sample gas flow | u_p | 0.077 | Vol.-% | 0.006 (Vol.-%) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 0.323 | Vol.-% | 0.105 (Vol.-%) ² |

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

| | | | |
|---|------------------------------------|------|--------|
| Combined standard uncertainty (u_c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 0.69 | Vol.-% |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 1.35 | Vol.-% |

Relative total expanded uncertainty

| | | |
|----------------------------------|------------------------------------|------------|
| Requirement of 2010/75/EU | U in % of the ELV 40 Vol.-% | 3.4 |
| Requirement of EN 15267-3 | U in % of the ELV 40 Vol.-% | 10.0 ** |
| | U in % of the ELV 40 Vol.-% | 7.5 |

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----|--------------------------|
| Certification range | HCl | 0 - 15 mg/m ³ |
|---------------------|-----|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|-------------------------|
| Sum of positive CS at zero point | 0.51 mg/m ³ |
| Sum of negative CS at zero point | 0.00 mg/m ³ |
| Sum of positive CS at span point | 0.51 mg/m ³ |
| Sum of negative CS at span point | -0.21 mg/m ³ |
| Maximum sum of cross-sensitivities | 0.51 mg/m ³ |
| Uncertainty of cross-sensitivity | 0.294 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u^2 |
|---|-----------|--------------------------|---|
| Repeatability standard deviation at set point * | u_r | 0.102 mg/m ³ | 0.010 (mg/m ³) ² |
| Lack of fit | u_{lof} | 0.063 mg/m ³ | 0.004 (mg/m ³) ² |
| Zero drift from field test | $u_{d,z}$ | -0.087 mg/m ³ | 0.008 (mg/m ³) ² |
| Span drift from field test | $u_{d,s}$ | 0.147 mg/m ³ | 0.022 (mg/m ³) ² |
| Influence of ambient temperature at span | u_t | 0.153 mg/m ³ | 0.023 (mg/m ³) ² |
| Influence of supply voltage | u_v | 0.083 mg/m ³ | 0.007 (mg/m ³) ² |
| Cross-sensitivity (interference) | u_i | 0.294 mg/m ³ | 0.087 (mg/m ³) ² |
| Influence of sample gas flow | u_p | 0.085 mg/m ³ | 0.007 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 0.121 mg/m ³ | 0.015 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

| | | |
|---|------------------------------------|------------------------|
| Combined standard uncertainty (u_c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 0.43 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 0.84 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|--|-------------|
| Requirement of 2010/75/EU | U in % of the ELV 10 mg/m³ | 8.4 |
| Requirement of EN 15267-3 | U in % of the ELV 10 mg/m³ | 40.0 |
| | U in % of the ELV 10 mg/m³ | 30.0 |

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|----|-------------------------|
| Certification range | HF | 0 - 3 mg/m ³ |
|---------------------|----|-------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.07 mg/m ³ |
| Sum of negative CS at zero point | -0.10 mg/m ³ |
| Sum of positive CS at span point | 0.04 mg/m ³ |
| Sum of negative CS at span point | 0.00 mg/m ³ |
| Maximum sum of cross-sensitivities | -0.10 mg/m ³ |
| Uncertainty of cross-sensitivity | -0.058 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u^2 |
|---|-----------|--------------------------|---|
| Repeatability standard deviation at set point * | u_r | 0.032 mg/m ³ | 0.001 (mg/m ³) ² |
| Lack of fit | u_{lof} | 0.017 mg/m ³ | 0.000 (mg/m ³) ² |
| Zero drift from field test | $u_{d,z}$ | -0.016 mg/m ³ | 0.000 (mg/m ³) ² |
| Span drift from field test | $u_{d,s}$ | 0.024 mg/m ³ | 0.001 (mg/m ³) ² |
| Influence of ambient temperature at span | u_t | 0.058 mg/m ³ | 0.003 (mg/m ³) ² |
| Influence of supply voltage | u_v | 0.012 mg/m ³ | 0.000 (mg/m ³) ² |
| Cross-sensitivity (interference) | u_i | -0.058 mg/m ³ | 0.003 (mg/m ³) ² |
| Influence of sample gas flow | u_p | 0.016 mg/m ³ | 0.000 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 0.024 mg/m ³ | 0.001 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

| | | |
|---|-----------------------------------|------------------------|
| Combined standard uncertainty (u_c) | $u_c = \sqrt{\sum (u_{max,i})^2}$ | 0.10 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 0.19 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|---|-------------|
| Requirement of 2010/75/EU | U in % of the ELV 1 mg/m³ | 19.4 |
| Requirement of EN 15267-3 | U in % of the ELV 1 mg/m³ | 40.0 |
| | U in % of the ELV 1 mg/m³ | 30.0 |

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|------------------|--------------------------|
| Certification range | N ₂ O | 0 - 50 mg/m ³ |
|---------------------|------------------|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|-------------------------|
| Sum of positive CS at zero point | 0.73 mg/m ³ |
| Sum of negative CS at zero point | 0.00 mg/m ³ |
| Sum of positive CS at span point | 1.50 mg/m ³ |
| Sum of negative CS at span point | -1.20 mg/m ³ |
| Maximum sum of cross-sensitivities | 1.50 mg/m ³ |
| Uncertainty of cross-sensitivity | 0.866 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u ² |
|--|------------------|-------------------------|---|
| Standard deviation from paired measurements under field conditions * | u _D | 0.171 mg/m ³ | 0.029 (mg/m ³) ² |
| Lack of fit | u _{lof} | 0.237 mg/m ³ | 0.056 (mg/m ³) ² |
| Zero drift from field test | u _{d,z} | 0.087 mg/m ³ | 0.008 (mg/m ³) ² |
| Span drift from field test | u _{d,s} | 0.404 mg/m ³ | 0.163 (mg/m ³) ² |
| Influence of ambient temperature at span | u _t | 0.400 mg/m ³ | 0.160 (mg/m ³) ² |
| Influence of supply voltage | u _v | 0.185 mg/m ³ | 0.034 (mg/m ³) ² |
| Cross-sensitivity (interference) | u _i | 0.866 mg/m ³ | 0.750 (mg/m ³) ² |
| Influence of sample gas flow | u _p | 0.162 mg/m ³ | 0.026 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u _{rm} | 0.404 mg/m ³ | 0.163 (mg/m ³) ² |

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

| | | |
|---|------------------------------------|------------------------|
| Combined standard uncertainty (u _c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 1.18 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 2.31 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|--|----------------|
| Requirement of 2010/75/EU | U in % of the ELV 50 mg/m³ | 4.6 |
| Requirement of EN 15267-3 | U in % of the ELV 50 mg/m³ | 20.0 ** |
| | U in % of the ELV 50 mg/m³ | 15.0 |

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----------------|--------------------------|
| Certification range | NH ₃ | 0 - 10 mg/m ³ |
|---------------------|-----------------|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.24 mg/m ³ |
| Sum of negative CS at zero point | -0.31 mg/m ³ |
| Sum of positive CS at span point | 0.08 mg/m ³ |
| Sum of negative CS at span point | -0.36 mg/m ³ |
| Maximum sum of cross-sensitivities | -0.36 mg/m ³ |
| Uncertainty of cross-sensitivity | -0.208 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u ² |
|---|------------------|--------------------------|---|
| Repeatability standard deviation at set point * | u _r | 0.115 mg/m ³ | 0.013 (mg/m ³) ² |
| Lack of fit | u _{lof} | 0.035 mg/m ³ | 0.001 (mg/m ³) ² |
| Zero drift from field test | u _{d,z} | -0.075 mg/m ³ | 0.006 (mg/m ³) ² |
| Span drift from field test | u _{d,s} | 0.069 mg/m ³ | 0.005 (mg/m ³) ² |
| Influence of ambient temperature at span | u _t | 0.153 mg/m ³ | 0.023 (mg/m ³) ² |
| Influence of supply voltage | u _v | 0.038 mg/m ³ | 0.001 (mg/m ³) ² |
| Cross-sensitivity (interference) | u _i | -0.208 mg/m ³ | 0.043 (mg/m ³) ² |
| Influence of sample gas flow | u _p | 0.037 mg/m ³ | 0.001 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u _{rm} | 0.081 mg/m ³ | 0.007 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

| | | |
|---|------------------------------------|------------------------|
| Combined standard uncertainty (u _c) | $u_c = \sqrt{\sum (u_{max, i})^2}$ | 0.32 mg/m ³ |
| Total expanded uncertainty | $U = u_c * k = u_c * 1.96$ | 0.62 mg/m ³ |

Relative total expanded uncertainty

| | | |
|---------------------------|--|----------------|
| Requirement of 2010/75/EU | U in % of the ELV 10 mg/m³ | 6.2 |
| Requirement of EN 15267-3 | U in % of the ELV 10 mg/m³ | 40.0 ** |
| | U in % of the ELV 10 mg/m³ | 30.0 |

** The EU-directive 2010/75/EU on industrial emissions provides no requirements for this component.
The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|----|---------------------------|
| Certification range | NO | 0 - 200 mg/m ³ |
|---------------------|----|---------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 1.64 mg/m ³ |
| Sum of negative CS at zero point | 0.00 mg/m ³ |
| Sum of positive CS at span point | 0.00 mg/m ³ |
| Sum of negative CS at span point | -6.30 mg/m ³ |
| Maximum sum of cross-sensitivities | -6.30 mg/m ³ |
| Uncertainty of cross-sensitivity | -3.637 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u^2 |
|--|-----------|--------------------------|--|
| Standard deviation from paired measurements under field conditions * | u_D | 0.819 mg/m ³ | 0.671 (mg/m ³) ² |
| Lack of fit | u_{lof} | 0.635 mg/m ³ | 0.403 (mg/m ³) ² |
| Zero drift from field test | $u_{d,z}$ | -0.231 mg/m ³ | 0.053 (mg/m ³) ² |
| Span drift from field test | $u_{d,s}$ | -1.155 mg/m ³ | 1.334 (mg/m ³) ² |
| Influence of ambient temperature at span | u_t | 1.249 mg/m ³ | 1.560 (mg/m ³) ² |
| Influence of supply voltage | u_v | 0.579 mg/m ³ | 0.335 (mg/m ³) ² |
| Cross-sensitivity (interference) | u_i | -3.637 mg/m ³ | 13.230 (mg/m ³) ² |
| Influence of sample gas flow | u_p | -0.818 mg/m ³ | 0.669 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u_{rm} | 1.617 mg/m ³ | 2.613 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max, i})^2} \quad 4.57 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 8.95 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 131 mg/m³ 6.8

Requirement of 2010/75/EU

U in % of the ELV 131 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 131 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----------------|--------------------------|
| Certification range | NO ₂ | 0 - 50 mg/m ³ |
|---------------------|-----------------|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.00 mg/m ³ |
| Sum of negative CS at zero point | 0.00 mg/m ³ |
| Sum of positive CS at span point | 0.50 mg/m ³ |
| Sum of negative CS at span point | -1.30 mg/m ³ |
| Maximum sum of cross-sensitivities | -1.30 mg/m ³ |
| Uncertainty of cross-sensitivity | -0.751 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u ² |
|--|------------------|--------------------------|---|
| Standard deviation from paired measurements under field conditions * | u _D | 0.111 mg/m ³ | 0.012 (mg/m ³) ² |
| Lack of fit | u _{lof} | 0.289 mg/m ³ | 0.084 (mg/m ³) ² |
| Zero drift from field test | u _{d,z} | 0.115 mg/m ³ | 0.013 (mg/m ³) ² |
| Span drift from field test | u _{d,s} | 0.462 mg/m ³ | 0.213 (mg/m ³) ² |
| Influence of ambient temperature at span | u _t | 0.208 mg/m ³ | 0.043 (mg/m ³) ² |
| Influence of supply voltage | u _v | 0.242 mg/m ³ | 0.059 (mg/m ³) ² |
| Cross-sensitivity (interference) | u _i | -0.751 mg/m ³ | 0.563 (mg/m ³) ² |
| Influence of sample gas flow | u _p | 0.235 mg/m ³ | 0.055 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u _{rm} | 0.404 mg/m ³ | 0.163 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 1.10 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 2.15 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 50 mg/m³ 4.3

Requirement of 2010/75/EU

U in % of the ELV 50 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

| | |
|-----------------------------------|-----------------------|
| Manufacturer | MKS Instruments Inc. |
| AMS designation | MGS300 |
| Serial number of units under test | 017151632 / 016842381 |
| Measuring principle | FTIR |

Test report

| | |
|-----------------|----------------|
| Test laboratory | 936/21208291/C |
| Date of report | TÜV Rheinland |
| | 2014-03-20 |

Measured component

| | | |
|---------------------|-----------------|--------------------------|
| Certification range | SO ₂ | 0 - 75 mg/m ³ |
|---------------------|-----------------|--------------------------|

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

| | |
|------------------------------------|--------------------------|
| Sum of positive CS at zero point | 0.71 mg/m ³ |
| Sum of negative CS at zero point | -1.76 mg/m ³ |
| Sum of positive CS at span point | 1.79 mg/m ³ |
| Sum of negative CS at span point | -2.09 mg/m ³ |
| Maximum sum of cross-sensitivities | -2.09 mg/m ³ |
| Uncertainty of cross-sensitivity | -1.208 mg/m ³ |

Calculation of the combined standard uncertainty

Tested parameter

| | | | u ² |
|--|------------------|--------------------------|---|
| Standard deviation from paired measurements under field conditions * | u _D | 0.348 mg/m ³ | 0.121 (mg/m ³) ² |
| Lack of fit | u _{lof} | 0.346 mg/m ³ | 0.120 (mg/m ³) ² |
| Zero drift from field test | u _{d,z} | -0.346 mg/m ³ | 0.120 (mg/m ³) ² |
| Span drift from field test | u _{d,s} | -0.606 mg/m ³ | 0.367 (mg/m ³) ² |
| Influence of ambient temperature at span | u _t | 0.643 mg/m ³ | 0.413 (mg/m ³) ² |
| Influence of supply voltage | u _v | 0.256 mg/m ³ | 0.066 (mg/m ³) ² |
| Cross-sensitivity (interference) | u _i | -1.208 mg/m ³ | 1.460 (mg/m ³) ² |
| Influence of sample gas flow | u _p | -0.352 mg/m ³ | 0.124 (mg/m ³) ² |
| Uncertainty of reference material at 70% of certification range | u _{rm} | 0.606 mg/m ³ | 0.368 (mg/m ³) ² |

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 1.78 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 3.48 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 50 mg/m³ 7.0

Requirement of 2010/75/EU

U in % of the ELV 50 mg/m³ 20.0

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 15.0