

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000072196

AMS designation: MGAprime Q for CO, NO, CO₂ and O₂

Manufacturer: MRU GmbH
Fuchshalde 8
74172 Neckarsulm
Germany

Test Laboratory: TÜV Rheinland Energy GmbH

**This is to certify that the P-AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-4 (2017),
EN 14793 (2017) and EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 9 pages).



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000072196

Publication in the German Federal Gazette
(BAnz) of 31 July 2020


This certificate will expire on:
30 July 2025

German Federal Environment Agency
Dessau, 07 September 2020

TÜV Rheinland Energy GmbH
Cologne, 06 September 2020



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.

Test Report:	936/21245785/B dated 11 May 2020
Initial certification:	31 July 2020
Expiry date:	30 July 2025
Publication:	BAnz AT 31.07.2020 B10, chapter I number 4.2

Approved application

The tested P-AMS is suitable for periodic measurements of stationary-source industrial emissions from plants requiring official approval in accordance with Directive 2010/75/EU (13th BImSchV), 44th BImSchV and TA Luft. As well it is suitable as an alternative method for the standard reference method for calibrating and validating stationary AMS in the context of QAL2 and AST in accordance with EN 14181. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the P-AMS for this application was assessed on the basis of a laboratory test and five field tests at different industrial plants. The plants were two municipal waste incinerators, a lignite-fired power plant, a sewage incinerator and a biomass heating plant.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the limit values and oxygen concentrations relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

Basis of the certification

This certification is based on:

- Test report no. 936/21245785/B dated 11 May 2020 issued by TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 31.07.2020 B10, chapter I number 4.2,
UBA announcement of 27 May 2020

AMS designation:

MGAprime Q for CO, NO, CO₂ and O₂

Manufacturer:

MRU GmbH, Neckarsulm

Field of application:

Portable AMS for periodic measurements of emissions from stationary sources in accordance with the 13th and 44th BImSchV, TA Luft and as an alternative method for the standard reference method for calibrating and validating stationary AMS in the context of QAL2 and AST in accordance with EN 14181

Measuring ranges during performance testing:

Component	Certification range	supplementary range	Unit
CO	0 – 220	0 – 3750	mg/m ³
NO	0 – 270	0 – 2680	mg/m ³
CO ₂	0–20	–	Vol.-%
O ₂	0–25	–	Vol.-%

Software version:

V1.001.029

Restrictions:

None

Notes:

1. Operation of the P-AMS requires that the pump provided by the manufacturer inject a 10% phosphoric acid be injected into the instrument's cooler.
2. In the event of temperature changes of more than 6 °C, it must be checked on-site whether the measurement uncertainty is still within the permissible limits.

Test Report:

TÜV Rheinland Energy GmbH, Cologne
Report no.: 936/21245785/B dated 11 May 2020

Certified product

This certification applies to automated measurement systems conforming to the following description:

The *MGAprime Q* measuring system under test consists of:

- HPI sampling probe (length: 0.3 m) and heating hose (length: 2 m),
- the *MGAprime-Q* analyser itself,
- and the additional APE unit for injecting phosphoric acid into the analyser.

The gas sampling probe has a heated probe handle incl. a dust filter and allows connection to a heated sampling line. The *MGAprime-Q* controls the heaters, the probe handle and the sampling probe.

The measuring system possesses two different kinds of sensors.

- a paramagnetic sensor to measure O₂;
- non-dispersive infrared absorption measurement of CO, CO₂ und NO.

The measuring system is operated via a touch-sensitive display. The outer casing of the instrument is a compact and robust metal housing with shock-absorbing plastic corners, housed in a water-repellent bag. The P-AMS can be operated from within this bag. The AMS has a degree of protection of IP42.

All electrical and pneumatic connections are located on the front of the unit.

An APE acid dosing unit is used to inject phosphoric acid (10%) into the instrument's cooler. Correct measurement requires the injection of phosphoric acid.

- This injection ensures constant conditions of the gas cooler.
- The use of phosphoric acid reduces losses of SO₂ and NO₂ on the wet surfaces of the gas cooler.

Two pumps ensure the injection. The total quantity of phosphorous acid during operation is 24ml per hour.

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at qal1.de.

Document history

Certification of the MGAprime Q measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

Initial certification according to EN 15267

Certificate no.0000072196: 07 September 2020
Expiry date of the certificate: 30 July 2025
Test report no. 936/21245785/B dated 11 May 2020
TÜV Rheinland Energy GmbH, Cologne
Publication: BANz AT 31.07.2020 B10, chapter I number 4.2
UBA announcement of 27 May 2020

The following pages present the uncertainty calculations for the individual components. Since separate uncertainty calculations are required for each field test, this certificate indicates the highest uncertainty result determined for each field test. All other uncertainty calculations are summarised in the mentioned performance test.

Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 1

Measuring system

Manufacturer	MRU GmbH
AMS designation	MGAprime Q
Serial number of units under test	63108 / 63107
Measuring principle	IR

Test report

Test laboratory	936/21245785/B TÜV Rheinland
Date of report	11 May 2020

Measured component

Certification range	CO 0 - 220 mg/m ³
---------------------	---------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.69 mg/m ³
Uncertainty of cross-sensitivity	u_i 0.399 mg/m ³

Calculation of the combined standard uncertainty

Test parameter

		u^2	
Standard deviation laboratory test	u_r 0.690 mg/m ³	0.476	(mg/m ³) ²
Lack of fit	u_{lof} 0.572 mg/m ³	0.327	(mg/m ³) ²
Zero drift from field test	$u_{d,7}$ 0.650 mg/m ³	0.423	(mg/m ³) ²
Span drift from field test	$u_{d,s}$ 1.299 mg/m ³	1.687	(mg/m ³) ²
Influence of ambient temperature from field	u_t 0.837 mg/m ³	0.700	(mg/m ³) ²
Influence of supply voltage field test specific	u_v 0.000 mg/m ³	0.000	(mg/m ³) ²
Cross-sensitivity field test specific	u_i 0.399 mg/m ³	0.159	(mg/m ³) ²
Influence of sample gas flow field test specific	u_b 0.000 mg/m ³	0.000	(mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm} 1.778 mg/m ³	3.162	(mg/m ³) ²

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	2.63 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	5.16 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 88 mg/m³	5.9
Requirement of EN 15267-3	U in % of the range 88 mg/m ³	7.5
Requirement for standard reference methods	U in % of the range 88 mg/m ³	6.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 3

Measuring system

Manufacturer	MRU GmbH
AMS designation	MGAprime Q
Serial number of units under test	63108 / 63107
Measuring principle	IR

Test report

Test laboratory	936/21245785/B TÜV Rheinland
Date of report	11.05.2020

Measured component

Certification range	NO 0 - 270 mg/m ³
---------------------	---------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	7.44 mg/m ³
Uncertainty of cross-sensitivity	u_i 4.297 mg/m ³

Calculation of the combined standard uncertainty

Test parameter

		u^2	
Standard deviation laboratory test	u_r 1.300 mg/m ³	1.690	(mg/m ³) ²
Lack of fit	u_{lof} 1.154 mg/m ³	1.332	(mg/m ³) ²
Zero drift from field test	$u_{d,7}$ 0.462 mg/m ³	0.213	(mg/m ³) ²
Span drift from field test	$u_{d,s}$ -1.386 mg/m ³	1.921	(mg/m ³) ²
Influence of ambient temperature from field	u_t 1.155 mg/m ³	1.333	(mg/m ³) ²
Influence of supply voltage field test specific	u_v 0.000 mg/m ³	0.000	(mg/m ³) ²
Cross-sensitivity field test specific	u_i 4.297 mg/m ³	18.464	(mg/m ³) ²
Influence of sample gas flow field test specific	u_b 0.000 mg/m ³	0.000	(mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm} 2.182 mg/m ³	4.763	(mg/m ³) ²

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	5.45 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	10.68 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 108 mg/m³	9.9
Requirement of EN 15267-3	U in % of the range 108 mg/m ³	15.0
Requirement for standard reference methods	U in % of the range 108 mg/m ³	10.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 1

Measuring system

Manufacturer	MRU GmbH
AMS designation	MGAprime Q
Serial number of units under test	63108 / 63107
Measuring principle	Paramagnetisch

Test report

Test laboratory	936/21245785/B TÜV Rheinland
Date of report	11.05.20

Measured component

Certification range	O ₂ 0 - 25 Vol.-%
---------------------	---------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.05 Vol.-%
Uncertainty of cross-sensitivity	u_i 0.030 Vol.-%

Calculation of the combined standard uncertainty

Test parameter

			u^2	
Standard deviation laboratory test	u_r	0.010 Vol.-%	0.000	(Vol.-%) ²
Lack of fit	u_{inf}	0.003 Vol.-%	0.000	(Vol.-%) ²
Zero drift from field test	$u_{d,z}$	0.017 Vol.-%	0.000	(Vol.-%) ²
Span drift from field test	$u_{d,s}$	-0.110 Vol.-%	0.012	(Vol.-%) ²
Influence of ambient temperature from field	u_t	0.245 Vol.-%	0.060	(Vol.-%) ²
Influence of supply voltage field test specific	u_v	0.000 Vol.-%	0.000	(Vol.-%) ²
Cross-sensitivity field test specific	u_i	0.030 Vol.-%	0.001	(Vol.-%) ²
Influence of sample gas flow field test specific	u_n	0.000 Vol.-%	0.000	(Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.202 Vol.-%	0.041	(Vol.-%) ²

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.34 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	0.66 Vol.-%

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 25 Vol.-%	2.6
Requirement of EN 15267-3	U in % of the range 25 Vol.-%	10.0 **
Requirement for standard reference methods	U in % of the range 25 Vol.-%	7.5
		6.0

** There are no requirements for this component in the EU Directive 2010/75/EU on industrial emissions. A value of 10.0 % was used.

Calculation of overall uncertainty according to EN 14181 and EN 15267-4 for both systems during field test 5

Measuring system

Manufacturer	MRU GmbH
AMS designation	MGAprime Q
Serial number of units under test	63108 / 63107
Measuring principle	IR

Test report

Test laboratory	936/21245785/B TÜV Rheinland
Date of report	11 May 2020

Measured component

Certification range	CO ₂ 0 - 20 Vol.-%
---------------------	----------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

The cross-sensitivities were calculated site-specifically as a function of the exhaust gas matrix at the respective field test facility, taking into account the cross-sensitivity influences determined in the laboratory.

Maximum sum of cross-sensitivities	0.16 Vol.-%
Uncertainty of cross-sensitivity	u_i 0.093 Vol.-%

Calculation of the combined standard uncertainty

Test parameter

			u^2	
Standard deviation laboratory test	u_r	0.020 Vol.-%	0.000	(Vol.-%) ²
Lack of fit	u_{inf}	0.115 Vol.-%	0.013	(Vol.-%) ²
Zero drift from field test	$u_{d,z}$	-0.012 Vol.-%	0.000	(Vol.-%) ²
Span drift from field test	$u_{d,s}$	-0.115 Vol.-%	0.013	(Vol.-%) ²
Influence of ambient temperature from field	u_t	0.245 Vol.-%	0.060	(Vol.-%) ²
Influence of supply voltage field test specific	u_v	0.000 Vol.-%	0.000	(Vol.-%) ²
Cross-sensitivity field test specific	u_i	0.093 Vol.-%	0.009	(Vol.-%) ²
Influence of sample gas flow field test specific	u_n	0.000 Vol.-%	0.000	(Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.162 Vol.-%	0.026	(Vol.-%) ²

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.35 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	0.68 Vol.-%

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 20 Vol.-%	3.4
Requirement of EN 15267-3	U in % of the range 20 Vol.-%	10.0 **
Requirement for standard reference methods	U in % of the range 20 Vol.-%	7.5
		6.0

** There are no requirements for this component in the EU Directive 2010/75/EU on industrial emissions. A value of 10.0 % was used.