

CONFIRMATION

Notification: 0000072196_01_04
on changes according to EN 15267 regarding certificate 0000072196_01 dated 2 June 2021

AMS designation: MGAprime Q for CO, NO, NO₂, N₂O, SO₂, O₂ and CO₂

Manufacturer: MRU GmbH
Fuchshalde 8
74172 Neckarsulm
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 August 2024
Publication: BAnz AT 31.10.2024 B9**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**25 Notification as regards Federal Environment Agency (UBA) notices
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter I number 4.1) and
of 5 July 2023 (BAnz AT 02.08.2023 B7, chapter III notification 13)**


The current software version for the portable measuring system MGAprime Q for CO, NO, NO₂, SO₂, N₂O, CO₂ and O₂ from the company MRU GmbH is: V1.003.009.


The sheath of the heated pipe can also be made of a more acid-resistant material (silicone foam).

The connection of the silicone hose to the pressure sensor for flow monitoring has been optimised with an adapter.

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 26 February 2024

TÜV Rheinland Energy & Environment GmbH
Cologne, 12 November 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.