

# CERTIFICATE

## of Product Conformity (QAL1)

**Certificate No. : 0000025927\_02**

**Certified AMS:** AR500 with ER120 for NO<sub>2</sub>, SO<sub>2</sub> and O<sub>3</sub>

**Manufacturer:** Opsis AB  
Skytteskogsvägen 16  
244 02 Furulund  
Sweden

**Test Institute:** TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested  
and found to comply with:**

**EN 14211: 2005, EN 14212: 2005, EN 14625: 2005,  
EN 15267-1: 2009 and EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate  
(see also the following pages).



Publication in the German Federal Gazette  
(BArz.) of 2 March 2012

This certificate will expire on:  
11 February 2020

German Federal Environment Agency  
Dessau, 2 February 2015

TÜV Rheinland Energie und Umwelt GmbH  
Cologne, 30 January 2015

i. A. Dr. Marcel Langner



ppa. Dr. Peter Wilbring



[www.umwelt-tuv.de](http://www.umwelt-tuv.de) / [www.eco-tuv.com](http://www.eco-tuv.com)  
teu@umwelt-tuv.de  
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH  
Am Grauen Stein  
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

**Certificate:**  
0000025927\_02 / 2 February 2015

**Test report:** 936/21211350/B of 7 October 2011

**Initial certification:** 12 February 2010

**Certificate:** renewal (previous certificate 0000025927\_01 of 16 March 2012  
valid until 11 February 2015)

**Date of expiry:** 11 February 2020

**Publication:** BAnz. 2 March 2012, no. 36, p. 920, chapter IV, no. 2.1

#### **Approved application**

The certified AMS is suitable for the continuous monitoring of concentrations of nitrogen dioxide, sulphur dioxide, and ozone in ambient air (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for the temperature range of +5 °C to +40 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

#### **Basis of the certification**

This certification is based on:

- test report 936/21211350/B dated 7 October 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 2 March 2012, no. 36, p. 920, chapter IV, No. 2.1, UBA announcement of 23 February 2012)
- publication in the German Federal Gazette (BAnz AT 5 March 2013 B10, chapter V, notice 11, UBA announcement of 12 February 2013)

**AMS designation:**AR500 with ER120 for NO<sub>2</sub>, SO<sub>2</sub> and O<sub>3</sub>**Manufacturer:**

Opsis AB, Furulund, Sweden

**Field of application:**

For the continuous monitoring of concentrations of nitrogen dioxide, sulphur dioxide and ozone in ambient air (stationary operation)

**Measuring ranges during the performance test:**

Component	Certification range	supplementary measurement ranges	Unit
NO <sub>2</sub>	0 - 400	0 - 1800	µg/m <sup>3</sup>
SO <sub>2</sub>	0 - 700	0 - 1000	µg/m <sup>3</sup>
O <sub>3</sub>	0 - 360	0 - 500	µg/m <sup>3</sup>

**Software version:**

7.21

**Restrictions:**

None

**Notes:**

1. The measuring path length during performance testing was 320 m.
2. The maintenance interval is four weeks.
3. Equivalence with the reference measurement methods according to the guideline „Demonstration of Equivalence of Ambient Air Monitoring Methods“ has been demonstrated for the components NO<sub>2</sub>, SO<sub>2</sub> and O<sub>3</sub>.
4. Functional tests by external sample gas feeding are possible.
5. The report on the performance test is available online at [www.qal1.de](http://www.qal1.de).
6. Supplementary testing (Demonstration of Equivalence for the component SO<sub>2</sub> according the guideline “Demonstration of Equivalence of Ambient Air Monitoring Methods”) as regards Federal Environment Agency notices of 25 January 2010 (BAnz. p. 552, chapter III no. 1.1).

**Test report:**TÜV Rheinland Energie und Umwelt GmbH, Cologne  
Report no.: 936/21211350/B dated 7 October 2011**11 Notification as regards Federal Environment Agency notices of 23 February 2012 (Federal Gazette (BAnz.) p. 920, chapter IV, no. 2.1)**

The AR500 measuring system with ER120 for NO<sub>2</sub>, SO<sub>2</sub> and O<sub>3</sub> manufactured by Op-sis AB can also be operated with the transmitter/receiver units ER 110 and ER150.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 10 October 2012

**Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The measurement system AR500 operates on the basis of the Differential Optical Absorption Spectroscopy (DOAS). The DOAS measuring principle uses the characteristic radiation absorption by gaseous components for quantification of the respective concentrations.

The DOAS monitor AR500 with ER120 consists of a combined transmitter-receiver unit, a reflector unit and an analyser. The absorbed light is transferred from the transmitter-receiver unit to the analyser via fibre optic cable.

**Combined transmitter-receiver Unit ER 120**

The combined transmitter-receiver unit ER120 comprises the optical components, the xenon light-source and the power supply PS150 for igniting the xenon light-source.

The used high-pressure xenon lamp is a point light source. The light is generated by ignition of ultra pure xenon gas at a pressure of approx. 30 bar. The lamp is powered by a stabilised D.C. voltage source and requires a short high-voltage ignition pulse.

The radiation of the lamp includes the ultraviolet, the visible, and the infrared range. The wavelengths are continuously distributed over the entire spectrum, with the exception of some peaks in the near infrared range.

**General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: [qal1.de](http://qal1.de).

Certification of the AR500 with ER120 for NO<sub>2</sub>, SO<sub>2</sub> and O<sub>3</sub> is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

**Initial certification according to EN 15267:**

Certificate No. 0000025927: 10 March 2010

Validity of the certificate: 11 February 2015

Test report: 936/21211350/A of 26 October 2009,  
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz. 12 February 2010, no. 24, p. 552, chapter III, no. 1.1  
UBA announcement of 25 January 2010

**Supplementary testing according to EN 15267:**

Certificate No. 0000025927\_01: 16 March 2012

Validity of the certificate: 11 February 2015

Test report: 936/21211350/B of 7 October 2011,  
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz. 2 March 2012, no. 36, p. 920, chapter IV, no. 2.1  
UBA announcement of 23 February 2012

**Notifications**

Statement of TÜV Rheinland Energie und Umwelt GmbH of 10 October 2012

Publication: BAnz AT 5 March 2013 B10, chapter V, notification 11 (additional transmitter/receiver)  
UBA announcement of 12 February 2013

**Renewal of the certificate**

Certificate No. 0000025927\_02 2 February 2015

Expiry date of the certificate: 11 February 2020

**Table 1:** Total expanded uncertainty with the results of the laboratory test according to EN 14211 (component NO<sub>2</sub>) for system 1329

Device: Component:	AR500 NO <sub>2</sub>	Serial No: 1329	104,6 nmol/mol		
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤	1,0 nmol/mol	0,000	0,0000
2	Repeatability at concentration ct	≤	3,0 nmol/mol	2,000	U <sub>t,z</sub> U <sub>t,fn</sub>
3	"lack of fit"	≤ 4,0% of measured value	0,800	U <sub>t,fn</sub>	0,48
4	Sensitivity coefficient of sample gas pressure	≤	8,0 nmol/mol/kPa	0,000	U <sub>gp</sub>
5	Sensitivity coefficient of sample gas temperature	≤	3,0 nmol/mol/K	0,026	U <sub>gt</sub>
6	Sensitivity coefficient of surrounding temperature	≤	3,0 nmol/mol/K	-0,050	U <sub>st</sub>
7	Sensitivity coefficient of electrical voltage	≤	0,30 nmol/mol/V	-0,021	U <sub>V</sub>
8a	H <sub>2</sub> O with concentration 21 nmol/mol	≤	5,0 nmol/mol	0,000	U <sub>H2O</sub>
8b	CO <sub>2</sub> with concentration 500 µmol/mol	≤	5,0 nmol/mol	0,001	U <sub>int, pos</sub>
8c	O <sub>3</sub> with concentration 200 nmol/mol	≤	2,0 nmol/mol	0,002	U <sub>int, neg</sub>
8d	NH <sub>3</sub> with concentration 200 nmol/mol	≤	5,0 nmol/mol	0,002	
9	Averaging effect	≤ 7,0% of measured value	-0,600	U <sub>av</sub>	-0,36
18	Difference sample/calibration port	≤ 1,0%	0,000	U <sub>asc</sub>	0,00
21	Converter efficiency	≥ 98%	100,000	U <sub>CE</sub>	0,00
22	Increase of NO <sub>2</sub> concentration due to residence time	≤ 4,0 nmol/mol	0,000	U <sub>ctr</sub>	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	U <sub>cg</sub>	1,05
combined standard uncertainty				U <sub>c</sub>	1,3046
expanded uncertainty				U <sub>c</sub>	2,6092
expanded uncertainty actual				U <sub>c,rel</sub>	<b>2,49</b>
expanded uncertainty required				U <sub>req,rel</sub>	15 %

**Table 2:** Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (component NO<sub>2</sub>) for system 1329

Device: Component:	AR500 NO <sub>2</sub>	Serial No: 1329	104,6 nmol/mol
No.	Performance characteristic	Criterion	Result
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	2,000
3	"lack of fit"	4,0% of measured value	0,800
4	Sensitivity coefficient of sample gas pressure	8,0 nmol/mol/kPa	0,000
5	Sensitivity coefficient of sample gas temperature	3,0 nmol/mol/K	0,026
6	Sensitivity coefficient of surrounding temperature	3,0 nmol/mol/K	-0,050
7	Sensitivity coefficient of electrical voltage	0,30 nmol/mol/V	-0,021
8a	H <sub>2</sub> O with concentration 21 nmol/mol	5,0 nmol/mol	0,000
8b	CO <sub>2</sub> with concentration 500 nmol/mol	5,0 nmol/mol	0,001
8c	O <sub>3</sub> with concentration 200 nmol/mol	2,0 nmol/mol	0,002
8d	NH <sub>3</sub> with concentration 200 nmol/mol	5,0 nmol/mol	0,002
9	Averaging effect	7,0% of measured value	-0,600
10	Reproducibility under field conditions	5,0% of the average of 3 Mon.	4,720
11	Long term drift at zero level	5,0 nmol/mol	-1,420
12	Long term drift at span level	5,0% of max. of certification range	0,430
18	Difference sample/calibration port	1,0%	0,000
21	Converter efficiency	98%	100,000
22	Increase of NO <sub>2</sub> concentration due to residence time	4,0 nmol/mol	0,000
23	Uncertainty calibration gas	3,0%	2,000
combined standard uncertainty		u <sub>c</sub>	7,1546
expanded uncertainty		U <sub>c</sub>	14,3093
expanded uncertainty actual		U <sub>c,rel</sub>	<b>13,68</b>
expanded uncertainty required		U <sub>req,rel</sub>	15 %

**Table 3:** Total expanded uncertainty with the results of the laboratory test according to EN 14211 (component NO<sub>2</sub>) for system 1330

No.	Performance characteristic	Criterion	Result	Uncertainty		Square of uncertainty
				1h-limit value:	1330	
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	$u_{r,Z}$	0,02	0,0003
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	$u_{r,h}$	0,02	0,0004
3	"lack of fit"	≤ 4,0% of measured value	0,000	$u_{l,h}$	0,36	1,1313
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	$u_{gp}$	0,00	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,000	$u_{gt}$	-0,05	0,0025
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	0,000	$u_{st}$	0,00	0,0000
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,000	$u_V$	0,24	0,0553
8a	H <sub>2</sub> O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	$u_{H2O}$	0,00	0,0000
8b	CO <sub>2</sub> with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,000	$u_{in, pos}$ or $u_{in, neg}$	0,42	0,1764
8c	O <sub>3</sub> with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,000			
8d	NH <sub>3</sub> with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,000			
9	Averaging effect	≤ 7,0% of measured value	0,000	$u_{av}$	-0,18	0,0328
18	Difference sample/calibration port	≤ 1,0%	0,000	$u_{isc}$	0,00	0,0000
21	Converter efficiency	≤ 98%	100,000	$u_{CE}$	0,00	0,0000
22	Increase of NO <sub>2</sub> concentration due to residence time	≤ 4,0 nmol/mol	0,000	$u_{ctr}$	0,00	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	$u_{cg}$	1,05	1,0941
combined standard uncertainty				$u_c$	1,2222	nmol/mol
expanded uncertainty				$U_c$	2,4445	nmol/mol
expanded uncertainty actual				$U_{c,rel}$	<b>2,34</b>	%
expanded uncertainty required				$U_{req,rel}$	15	%

**Table 4:** Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (component NO<sub>2</sub>) for system 1330

Device: Component:	AR500 NO <sub>2</sub>	Serial No.: 1330	104,6 nmol/mol		
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	$U_{t,Z}$ $U_{r,h}$	0,02 0,0003
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	$U_{r,h}$	not respected because, $U_{r,h} = 0,034 < U_{r,f}$ -
3	"lack of fit"	≤ 4,0% of measured value	0,600	$U_{t,h}$	0,36 0,1313
4	Sensitivity coefficient of sample gas pressure	8,0 nmol/mol/kPa	0,000	$U_{gp}$	0,00 0,0000
5	Sensitivity coefficient of sample gas temperature	3,0 nmol/mol/K	-0,032	$U_{gt}$	-0,05 0,0025
6	Sensitivity coefficient of surrounding temperature	3,0 nmol/mol/K	0,000	$U_{st}$	0,00 0,0000
7	Sensitivity coefficient of electrical voltage	0,30 nmol/mol/V	0,073	$U_V$	0,24 0,0553
8a	H2O with concentration 21 mmol/mol	5,0 nmol/mol	0,000	$U_{H2O}$	0,00 0,0000
8b	CO2 with concentration 500 µmol/mol	5,0 nmol/mol	0,001	$U_{in, pos}$	0,1764
8c	O3 with concentration 200 nmol/mol	2,0 nmol/mol	0,002	or $U_{in, neg}$	0,42 0,0000
8d	NH3 with concentration 200 nmol/mol	5,0 nmol/mol	0,000		
9	Averaging effect	≤ 7,0% of measured value	-0,300	$U_{av}$	-0,18 0,0328
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,720	$U_{f,f}$	4,94 24,3752
11	Long term drift at zero level	≤ 5,0 nmol/mol	1,620	$U_{d,LZ}$	0,94 0,8748
12	Long term drift at span level	≤ 5,0% of max. of certification range	0,500	$U_{d,U,h}$	0,30 0,0912
18	Difference sample/calibration port	1,0%	0,000	$U_{DSc}$	0,00 0,0000
21	Converter efficiency	0,98	100,000	$U_{CE}$	0,00 0,0000
22	Increase of NO <sub>2</sub> concentration due to residence time	4,0 nmol/mol	0,000	$U_{ctr}$	0,00 0,0000
23	Uncertainty calibration gas	3,0%	2,000	$U_{cg}$	1,05 1,0941
		combined standard uncertainty	$U_c$	7,1561 expanded uncertainty	nmol/mol nmol/mol
			$U_c$	14,3121 expanded uncertainty actual	%
			$U_{c,rel}$	13,68 expanded uncertainty required	%
			$U_{req,rel}$	15	%

**Certificate:**  
0000025927\_02 / 2 February 2015

**Table 5:** Total expanded uncertainty with the results of the laboratory test according to EN 14212 (component SO<sub>2</sub>) for system 1329

No.	Device: AR500  Component: SO <sub>2</sub>	Serial-No.: Geraet 1 (1329)			
		1h-limit value: nmol/mol	132	132	nmol/mol
1	Performance characteristic Repeatability at zero	≤ 1,0 nmol/mol	0,100	$U_{1,Z}$	0,02
2	Repeatability at concentration ct "lack of fit"	≤ 3,0 nmol/mol	0,100	$U_{1,V}$	0,02
3	Sensitivity coefficient of sample gas pressure	≤ 4,0% of measured value	1,600	$U_{1,V}$	1,22
4	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/kPa	0,000	$U_{gp}$	0,00
5	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	0,071	$U_{gr}$	0,54
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	$U_{st}$	-0,23
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	$U_V$	-0,10
8a	H <sub>2</sub> O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	$U_{H2O}$	0,00
8b	H <sub>2</sub> S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	$U_{int, pos}$	
8c	NH <sub>3</sub> with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406	or	
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604		0,1600
8e	NO <sub>2</sub> with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404		
8f	m-Xylo with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421	$U_{int,neg}$	
9	Averaging effect	≤ 7,0% of measured value	-0,100	$U_{av}$	-0,08
18	Difference sample/calibration port	≤ 1,0%	0,000	$U_{psc}$	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	$ug$	1,7424
				combined standard uncertainty	1,9363
				expanded uncertainty	3,8726
				expanded uncertainty actual	<b>2,93</b>
				expanded uncertainty required	15 %

**Table 6:** Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (component SO<sub>2</sub>) for system 1329

Device: Component:	AR500 SO <sub>2</sub>	Serial-No.: 132 nmol/mol			
		1h-limit value:	Result	Uncertainty	Square of uncertainty
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	$u_{t,z}$	0,02
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	$u_{t,v}$	not respected because, $u_r/v = 0,01 < u_{t,f}$
3	"lack of fit"	≤ 4,0% of measured value	1,600	$u_{t,v}$	1,22
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	$u_{g,p}$	0,000
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,071	$u_{g,t}$	0,54
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	$u_{s,t}$	-0,23
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	$u_V$	-0,10
8a	H <sub>2</sub> O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	$u_{H_2O}$	0,000
8b	H <sub>2</sub> S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	$u_{int, pos}$	
8c	NH <sub>3</sub> with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406		0,1600
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604	or	
8e	NO <sub>2</sub> with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404		
8f	m-Xylool with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421	$u_{int,neg}$	
9	Averaging effect	≤ 7,0% of measured value	-0,100	$u_{av}$	-0,08
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,830	$u_{t,f}$	6,38
11	Long term drift at zero level	≤ 5,0 nmol/mol	-0,920	$u_{d,l,z}$	-0,53
12	Long term drift at span level	≤ 5,0% of max. of certification range	1,490	$u_{d,l,v}$	1,14
18	Differenz Proben-/Kalibriergaseingang	≤ 1,0%	0,000	$u_{bias}$	0,000
23	Unsicherheit Prüfgas	≤ 3,0%	2,000	$u_{cg}$	1,32
			combined standard uncertainty	$u_c$	6,7800
			expanded uncertainty	$U_{c,[el]}$	13,5600
			expanded uncertainty actual	$U_{c,[el]}$	10,27
			expanded uncertainty required	$U_{eq,[el]}$	15 %

**Table 7:** Total expanded uncertainty with the results of the laboratory test according to EN 14212 (component SO<sub>2</sub>) for system 1330

No.	Performance characteristic	Criterion	Result	Serial-No.: Gerät 2 (1330)		1h-limit value: 132 nmol/mol
				Uncertainty	Square of uncertainty	
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	$U_{r,Z}$	0,00	0,0000
2	Repeatability at concentration ct "lack of fit"	≤ 3,0 nmol/mol	0,100	$U_{r,lv}$	0,02	0,0003
3	Sensitivity coefficient of sample gas pressure	4,0% of measured value	1,400	$U_{l,lv}$	1,07	1,1384
4	Sensitivity coefficient of sample gas temperature	3,0 nmol/mol/kPa	0,000	$U_{gp}$	0,00	0,0000
5	Sensitivity coefficient of surrounding temperature	1,0 nmol/mol/K	0,011	$U_{gr}$	0,08	0,0070
6	Sensitivity coefficient of surrounding temperature	1,0 nmol/mol/K	-0,060	$U_{st}$	-0,46	0,2091
7	Sensitivity coefficient of electrical voltage	0,30 nmol/mol/V	0,010	$U_V$	0,10	0,0103
8a	H <sub>2</sub> O with concentration 21 nmol/mol	10 nmol/mol	0,000	$U_{H2O}$	0,00	0,0000
8b	H <sub>2</sub> S with concentration 200 nmol/mol	5,0 nmol/mol	0,503	$U_{int, pos}$		
8c	NH <sub>3</sub> with concentration 200 nmol/mol	5,0 nmol/mol	0,203			
8d	NO with concentration 500 nmol/mol	5,0 nmol/mol	0,202	or	1,23	1,5129
8e	NO <sub>2</sub> with concentration 200 nmol/mol	5,0 nmol/mol	0,401			
8f	m-Xylo with concentration 1 μmol/mol	10 nmol/mol	0,809	$U_{int, neg}$		
9	Averaging effect	7,0% of measured value	0,100	$U_{av}$	0,08	0,0058
18	Difference sample/calibration port	1,0%	0,000	$U_{Dsc}$	0,00	0,0000
23	Uncertainty calibration gas	3,0%	2,000	0	1,32	1,7424
				combined standard uncertainty	$U_c$	2,1509 nmol/mol
				expanded uncertainty	$U_c$	4,3017 nmol/mol
				expanded uncertainty actual	$U_{c,rel}$	3,26 %
				expanded uncertainty required	$U_{req,rel}$	15 %

**Table 8:** Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (component SO<sub>2</sub>) for system 1330

Device: Component:	AR500 SO <sub>2</sub>	Serial-No.: Gerät 2 (1330)			
		1h-limit value: nmol/mol	132	132	nmol/mol
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	u <sub>i,z</sub> not respected because, u <sub>i,v</sub> = 0,01 < u <sub>r,f</sub>	0,0000
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	u <sub>i,v</sub>	-
3	"lack of fit"	≤ 4,0% of measured value	1,400	u <sub>i,v</sub>	1,1384
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/K <sup>a</sup>	0,000	u <sub>gp</sub>	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,011	u <sub>gt</sub>	0,0070
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,060	u <sub>st</sub>	0,2091
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,010	u <sub>v</sub>	0,0103
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	u <sub>H2O</sub>	0,0000
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,503	u <sub>Hn, pos</sub>	-
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,203		-
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	0,202	or	1,5129
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,401		-
8f	m-Xylool with concentration 1 µmol/mol	≤ 10 nmol/mol	0,809	u <sub>In, neg</sub>	-
9	Averaging effect	≤ 7,0% of measured value	0,100	u <sub>av</sub>	0,0058
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,830	u <sub>r,f</sub>	40,6483
11	Long term drift at zero level	≤ 5,0 nmol/mol	1,160	u <sub>d,l,z</sub>	0,67
12	Long term drift at span level	≤ 5,0% of max. of certification range	-2,070	u <sub>d,l,v</sub>	2,4887
18	Differenz Proben-/Kalibriergasgang	≤ 1,0%	0,000	u <sub>Dsc</sub>	0,0000
23	Unsicherheit Prüfgas	≤ 3,0%	2,000	0	1,7424
		combined standard uncertainty			
		expanded uncertainty			
		expanded uncertainty actual			
		expanded uncertainty required			

**Table 9:** Total expanded uncertainty with the results of the laboratory test according to EN 14625 (component O<sub>3</sub>) for system 1329

Measured component:	Device: AR500  O <sub>3</sub>	Serial No.	Gefäß 1 (1329)		
			hourly alert threshold 120 nmol/mol	hourly alert threshold 120 nmol/mol	hourly alert threshold 120 nmol/mol
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	U <sub>t,Z</sub> 0,04	0,0013
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	U <sub>t,IV</sub> 0,11	0,0120
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	U <sub>t,IV</sub> 0,28	0,0768
4	Variations in sample gas pressure	2,0 nmol/mmol/kPa	0,000	U <sub>gp</sub> 0,00	0,0000
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	U <sub>gt</sub> 0,15	0,0212
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	U <sub>st</sub> 0,52	0,2700
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	U <sub>V</sub> -0,12	0,0147
8a	Interference H <sub>2</sub> O with 21 nmol/mol	10 nmol/mol	0,000	U <sub>H2O</sub> 0,00	0,0000
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	U <sub>int, pos</sub> 1,47	2,1573
8c	Interference Xylool with 0,5 µmol/mol	5,0 nmol/mol	0,397 or U <sub>int, neg</sub>		
9	Averaging effect	7,0% of measured value	0,200	U <sub>av</sub> 0,14	0,0192
18	Difference sample/calibration port	1,0%	0,000	U <sub>Dsc</sub> 0,00	0,0000
23	Uncertainty test gas	3,0%	2,000	U <sub>CG</sub> 1,20	1,4400
		Combined standard uncertainty	U <sub>c</sub>	2,0031 nmol/mol	nmol/mol
		Expanded uncertainty	U <sub>c</sub>	4,0062 3,34	nmol/mol %
		Expanded uncertainty actual	U <sub>c,rel</sub>		
		Expanded uncertainty required	U <sub>req,rel</sub>	15	%

**Table 10**

Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (component O<sub>3</sub>) for system 1329

Measured component	Device: AR500	O <sub>3</sub>	Serie No. Gefäß 1 (1329)			
			hourly alert threshold nmol/mol	120 nmol/mol	Result	Uncertainty
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty	
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	U <sub>t,z</sub> U <sub>t,lv</sub>	0,04	0,0013
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	U <sub>t,lv</sub>	not respected because, u <sub>t,lv</sub> = 0,1 < u <sub>t,f</sub>	-
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	U <sub>t,lv</sub>	0,28	0,0768
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	U <sub>gp</sub>	0,00	0,0000
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	U <sub>gt</sub>	0,15	0,0212
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	U <sub>st</sub>	0,52	0,2700
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	U <sub>v</sub>	-0,12	0,0147
8a	Interference H2O with 2,1 nmol/mol	10 nmol/mol	0,000	U <sub>H2O</sub>	0,00	0,0000
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	U <sub>int, pos</sub> or U <sub>int, neg</sub>	1,47	2,1573
8c	Interference Xylool with 0,5 µmol/mol	5,0 nmol/mol	0,397			
9	Averaging effect	7,0% of measured value	0,200	U <sub>av</sub>	0,14	0,0192
10	Reproducibility standard deviation in field	5,0% of average of 3 month	2,410	U <sub>t,f</sub>	2,89	8,3637
11	Long term drift at zero	5,0 nmol/mol	1,460	U <sub>d,lz</sub>	0,84	0,7105
12	Long term drift at span level	5,0% of range	-2,450	U <sub>d,lv</sub>	-1,70	2,8812
18	Difference sample/calibration port	1,0%	0,000	uDSC	0,00	0,0000
23	Uncertainty test gas	3,0%	2,000	ug	1,20	1,4400
				Combined standard uncertainty	U <sub>c</sub>	3,9945
				Expanded uncertainty	U <sub>c</sub>	7,9890
				Expanded uncertainty actual	U <sub>c,rel</sub>	<b>6,66</b>
				Expanded uncertainty required	U <sub>req,el</sub>	15 %

**Table 11:** Total expanded uncertainty with the results of the laboratory test according to EN 14625 (component O<sub>3</sub>) for system 1330

Measured component:		AR500		O <sub>3</sub>		hourly alert threshold		120		nmol/mol	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty						
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	U <sub>i,Z</sub>	0,04						0,0013
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,400	U <sub>i,V</sub>	0,07						0,0053
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	-0,300	U <sub>i,V</sub>	-0,21						0,0432
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	U <sub>gp</sub>	0,00						0,0000
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,007	U <sub>gt</sub>	0,07						0,0053
6	Variations in surrounding temperature	1,0 nmol/mol/K	-0,120	U <sub>st</sub>	-0,42						0,1728
7	Variations in electrical voltage	0,30 nmol/mol/V	0,010	U <sub>v</sub>	0,12						0,0147
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	U <sub>H2O</sub>	0,00						0,0000
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	0,396	U <sub>int, pos</sub>	1,72						2,9416
8c	Interference Xylool with 0,5 µmol/mol	5,0 nmol/mol	2,574	or U <sub>int, neg</sub>							
9	Averaging effect	7,0% of measured value	-0,900	U <sub>av</sub>	-0,62						0,3888
18	Difference sample/calibration port	1,0%	0,000	U <sub>Dsc</sub>	0,00						0,0000
23	Uncertainty test gas	3,0%	2,000	0	1,20						1,4400
				Combined standard uncertainty	U <sub>c</sub>						2,2390
				Expanded uncertainty	U <sub>c</sub>						4,4780
				Expanded uncertainty actual	U <sub>c,rel</sub>						3,73
				Expanded uncertainty required	U <sub>req,rel</sub>						15 %

**Table 12** Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (component O<sub>3</sub>) for system 1330

No.	Measured component:	Device: AR500	O <sub>3</sub>	hourly alert threshold				Series No. Gerät 2 (1330)	120 nmol/mol
				Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability standard deviation at zero		1,0 nmol/mol	0,200	u <sub>r,z</sub>	0,04	0,0013		
2	Repeatability standard deviation at ct		3,0 nmol/mol	0,400	u <sub>r,lv</sub>				
3	"lack of fit" at the hourly alert threshold value		4,0% of measured value	-0,300	u <sub>l,lv</sub>	-0,21	0,0432		
4	Variations in sample gas pressure		2,0 nmol/mol/kPa	0,000	u <sub>gp</sub>	0,00	0,0000		
5	Variations in sample gas temperature		1,0 nmol/mol/K	0,007	u <sub>gt</sub>	0,07	0,0053		
6	Variations in surrounding temperature		1,0 nmol/mol/K	-0,120	u <sub>st</sub>	-0,42	0,1728		
7	Variations in electrical voltage		0,30 nmol/mol/V	0,010	u <sub>v</sub>	0,12	0,0147		
8a	Interference H2O with 21 nmol/mol		10 nmol/mol	0,000	u <sub>H2O</sub>	0,00	0,0000		
8b	Interference Toloul with 0,5 µmol/mol		5,0 nmol/mol	0,396	u <sub>mt, pos</sub> or u <sub>mt, neg</sub>	1,72	2,9416		
8c	Interference Xylool with 0,5 µmol/mol		5,0 nmol/mol	2,574					
9	Averaging effect		7,0% of measured value	-0,900	u <sub>av</sub>	-0,62	0,3888		
10	Reproducibility standard deviation in field		5,0% of average of 3 month	2,410	u <sub>rf</sub>	2,89	8,3637		
11	Long term drift at zero		5,0 nmol/mol	-1,840	u <sub>dl,z</sub>	-1,06	1,1285		
12	Long term drift at span level		5,0% of range	2,900	u <sub>dl,lv</sub>	2,01	4,0368		
18	Difference sample/calibration point		1,0%	0,000	uDsc	0,00	0,0000		
23	Uncertainty test gas		3,0%	2,000	0	1,20	1,4400		
			Combined standard uncertainty		u <sub>c</sub>		4,3054	nmol/mol	
			Expanded uncertainty		u <sub>c</sub>		8,6109	nmol/mol	
			Expanded uncertainty actual		u <sub>c,rel</sub>		7,18	%	
			Expanded uncertainty required		u <sub>req,rel</sub>		15	%	