

# CONFIRMATION

Notification: 0000040212\_06\_01  
on changes according to EN 15267 regarding certificate 0000040212\_06 dated 20 March 2024

---

**AMS designation:** Fidas 200 S, Fidas 200, Fidas 200 E for PM<sub>10</sub> and PM<sub>2,5</sub>

**Manufacturer:** PALAS GmbH  
Greschbachstrasse 3b  
76229 Karlsruhe  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 19 March 2024  
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**57 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and  
of 5 July 2023 (BAnz AT 02.08.2023 B7, chapter III notification 36)**

The current software version for the measuring systems Fidas® 200 S, Fidas® 200 E  
and Fidas® 200 for PM<sub>10</sub> and PM<sub>2,5</sub> from Palas GmbH is

100539.0014.0001.0001.0011


Firmware 400045\_FW\_V0314 can also be used for the compact weather station from  
the manufacturer Thies.


The circuit board layout of the MMS-PT100 has changed to Rev. 7.1

Statement issued by TÜV Rheinland Energy GmbH dated 01 September 2023

---

TÜV Rheinland Energy & Environment GmbH  
Cologne, 24 May 2024

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.