

CONFIRMATION

Notification: 0000040212_06_04
on changes according to EN 15267 regarding certificate 0000040212_06 dated 20 March 2024

AMS designation: Fidas® 200 S, Fidas® 200, Fidas® 200 E for PM₁₀ and PM_{2,5}

Manufacturer: PALAS GmbH
Siemensallee 84 / Gebäude 7330
76187 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 27 August 2025
Publication: BAnz AT 31.10.2025 B5**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**42 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 5.1) and
of 2 April 2025 (BAnz AT 19.05.2025 B3, chapter IV notification 73)**

The current software version for the Fidas® 200 S, Fidas® 200 E
and Fidas® 200 for PM₁₀ and PM_{2,5} from Palas GmbH is:


100544.0014.0001.0001.0011


This includes version 100543.0014.0001.0001.0011.

The flow sensor (Siargo FS4008-10-O6-CV-A) for the Fidas® 200 S, Fidas® 200 E and
Fidas® 200 measuring devices is being replaced by the successor model (Siargo
FS4308-10-O6-BV-A-0.5/4.5V).

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 May 2025

TÜV Rheinland Energy & Environment GmbH
Cologne, 21 November 2025


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
qa11-info@tuv.com
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.