

CERTIFICATE

of Product Conformity (QAL1)

Certificate No. : 0000025930_03

Certified AMS: Modular System MAC GMS800 for CO, NO, NO₂, SO₂, CH₄, N₂O, CO₂ and O₂

Manufacturer: SICK AG
Nimburger Str. 11
79276 Reute
Germany

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2008
and EN 14181: 2004**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000025930

Publication in the German Federal Gazette
(BAnz.) of 2 March 2012


German Federal Environment Agency
Dessau, 2 February 2015



i. A. Dr. Marcel Langner

This certificate will expire on:
11 February 2020

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 30 January 2015



ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

Test report:	936/21217568/A of 18 October 2011
Initial certification:	12 February 2010
Certificate:	renewal (previous certificate 0000025930_02 of 16 March 2012 valid until 11 February 2015)
Expiry date:	11 February 2020
Publication:	BAnz. 2 March 2012, no. 36, p. 920, chapter I, no. 5.1

Approved application

The tested AMS is suitable for use at large combustion plants according to Directive 2001/80/EC, at waste incineration plants according to Directive 2000/76/EC and other plants requiring official approval. The tested ranges have been chosen with respect to the wide application of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a field test at a municipal waste incineration plant.

The AMS is approved for an ambient temperature range of +5 °C to +40 °C.

The notification of suitability of the AMS, performance testing, and the uncertainty calculation have been effected on the basis of the regulations valid at the time of performance testing.

Any potential user should ensure that this AMS is suitable for monitoring the limit values relevant to the application.

Basis of the certification

This certification is based on:

- test report 936/21217568/A dated 18 October 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 2 March 2012, no. 36, p. 920, chapter I, no. 5.1, UBA announcement of 23 February 2012)
- publication in the German Federal Gazette (BAnz AT 23 July 2013 B4, chapter V, notifications 12 (sequential no. 12) and 13 (sequential no. 13), UBA announcement of 3 July 2013)
- publication in the German Federal Gazette (BAnz AT 5 August 2014 B11, chapter V, notification 13, UBA announcement of 17 July 2014)

AMS designation:

Modular System MAC GMS800 for CO, NO, NO₂, SO₂, CH₄, N₂O, CO₂ and O₂

Manufacturer:

SICK MAIHAK GmbH, Reute

Field of application:

For measurements at plants requiring official approval (i.e. Directive 2001/80/EC regarding large combustions plants, Directive 2000/76/EC regarding waste incineration plants)

Measuring ranges during the performance test:

Component	Module	Certification range	Supplementary measuring ranges		Unit
CO	MAC GMS800 UNOR for CO	0 – 75	0 – 750	0 – 3000	mg/m ³
	MAC GMS800 MULTOR for CO	0 – 200	0 – 2000	–	mg/m ³
NO	MAC GMS800 UNOR for NO	0 – 100	0 – 1000	0 – 2000	mg/m ³
	MAC GMS800 MULTOR for NO	0 – 250	0 – 2500	–	mg/m ³
	MAC GMS800 DEFOR for NO	0 – 50	0 – 1000	0 – 2000	mg/m ³
NO ₂	MAC GMS800 DEFOR for NO ₂	0 – 50	0 – 500	–	mg/m ³
NO _x	MAC GMS800 UNOR for NO _x	0 – 100	0 – 1000	0 – 2000	mg/m ³
	MAC GMS800 MULTOR for NO _x	0 – 250	0 – 2500	–	mg/m ³
SO ₂	MAC GMS800 UNOR for SO ₂	0 – 75	0 – 287	0 – 2000	mg/m ³
	MAC GMS800 MULTOR for SO ₂	0 – 250	0 – 2000	–	mg/m ³
	MAC GMS800 DEFOR for SO ₂	0 – 75	0 – 287	0 – 2000	mg/m ³
CH ₄	MAC GMS800 UNOR for CH ₄	0 – 50	0 – 500	–	mg/m ³
	MAC GMS800 MULTOR for CH ₄	0 – 286	0 – 500	–	mg/m ³
N ₂ O	MAC GMS800 UNOR for N ₂ O	0 – 50	0 – 500	–	mg/m ³
CO ₂	MAC GMS800 UNOR for CO ₂	0 – 25	–	–	Vol.-%
	MAC GMS800 MULTOR for CO ₂	0 – 25	–	–	Vol.-%
O ₂	MAC GMS800 OXOR-P for O ₂	0 – 25	–	–	Vol.-%
	MAC GMS800 OXOR-E for O ₂	0 – 25	–	–	Vol.-%

Software versions:

T825_090707_1000

PC-Software: Sopas ET 2.22 Build 2938

Restrictions:

1. Functionality of the respective combination of modules shall be verified during the checks for proper installation.
2. The maintenance interval shall be determined during the check for proper installation.

Notes:

1. Automatic calibration of zero points shall be carried out with humidified ambient air for all components except for O₂ (OXOR-P and OXOR-E) once a week.
2. Automatic span point calibration for the OXOR-P and OXOR-E (O₂) sensors shall be carried out once a week with humidified ambient air.
3. With the help of external air conditioning the AMS also fulfils the requirements at an ambient air temperature of 50 °C.
4. The measuring system may be operated with cooler type MAK10-2 manufactured by AGT Thermotechnik as well as with type CSS-V2SK manufactured by M&C.
5. With weekly adjustments using the respective internal test gas cell or edge filter (NO₂ (DEFOR)), the maintenance intervals for the modules can be extended as follows:
 - one year for the modules CO (UNOR), CH₄ (UNOR and MULTOR)
 - half a year for the modules CO (MULTOR), NO (MULTOR), SO₂ (DEFOR)
 - three months for the modules NO (UNOR) und NO₂ (DEFOR)
6. Supplementary testing (extension of the maintenance interval by using internal test gas cells) as regards Federal Environment Agency notices of 12 July 2010 (BAnz. p. 2597, chapter I, no. 2.1) and of 10 January 2011 (BAnz. p. 294, chapter IV notifications 2 and 30).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21217568/A of 18 October 2011

12 Notification as regards Federal Environment Agency notices regarding performance tested measuring systems manufactured by SICK MAIHAK GmbH

Seq. no.	AMS / Manufacturer	Notice	Notification	Statement of test institute
12	MAC GMS800 / SICK AG	of 23 February 2012 (BAnz. p. 920, chapter I no. 5.1)	SICK MAIHAK GmbH merged with its parent company SICK AG as of 1 January 2013. The manufacturer is now registered as SICK AG.	Statement of TÜV Rheinland Energie und Umwelt GmbH of 25 März 2013

13 Notification as regards Federal Environment Agency notices regarding performance tested measuring systems manufactured by SICK Engineering GmbH and SICK AG

Seq. no.	AMS / Manufacturer	Notice	Notification	Statement of test institute
13	MAC GMS800 / SICK AG	as regards notification 12 (sequential no. 12) of this notice	The current software version of the SOPAS ET platform for optional AMS control is: SOPAS ET 2.38.	Statement of TÜV Rheinland Energie und Umwelt GmbH of 25 March 2013

13 Notification as regards Federal Environment Agency notices of 23 February 2012 (BAnz. p. 920, chapter 1, no. 5.1), of 3 July 2013 (BAnz AT 23 July 2013 B4, chapter V, 12th notification [no.12] and 13th notification [no. 13]) and of 27 February 2014 (BAnz AT 1 April 2014 B12, chapter V, 1st correction)

The modular measuring system MAC GMS800 for CO, NO, NO₂, SO₂, CH₄, N₂O, CO₂ and O₂ manufactured by SICK AG may now also be equipped with the SCU-P100 display unit.

For the DEFOR module, an absorber cartridge is inserted into the measurement cell.

The chopper motor S/N 6026930 is replaced by motor S/N 6030437.

The software versions for the individual modules of the MAC GMS800 modular measuring system for CO, NO, NO₂, SO₂, CH₄, N₂O, CO₂ and O₂ manufactured by SICK AG are:

BCU: 9150883_3.005 Y123
 SCU-P100: 9158931_WI82
 UNOR/MULTOR: 9137995_3.004 XN94
 OXOR: 9138052_3.002 WM48
 DEFOR: 9139736_3.003 WM48
 Gas module: 9137582_3.002 WM48

Statement of TÜV Rheinland Energie und Umwelt GmbH of 2 April 2014

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The multi-component measuring system MAC GMS800 is a modular sensor system. The essential part is the instrument cabinet including the interface modules, measuring gas pump, test gas supply unit, electronic-unit and SCU/BCU control unit. It is possible to place up to three different measurement modules in this instrument cabinet. All gas sensors are able to work independently from other sensors.

Thus, the modular measurement system can be equipped according to different requirements, each with appropriate measurement modules.

The following gas sensor modules have been certified so far: UNOR, MULTOR, DEFOR, OXOR.

All gas sensor modules are controlled by a BUS-system. The data output and adjustment of all sensors can be observed with this system.

The following components are part of the complete system:

- heated probe (M&C SP 2000) with heated filter, test gas offering function and back-flush function,
- heated gas tube (a heated line with a length of 10 m was used during the laboratory investigations, during field investigations a heated line with a length of 50 m was used),
- instrument cabinet with interface modules, measuring gas pump, sample gas cooler, test gas supply unit, sensor modules with gas sensors, electronic-unit and SCU/BCU control unit.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of the modular system MAC GMS800 for CO, NO, NO₂, SO₂, CH₄, N₂O, CO₂ and O₂ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267:

Certificate No.0000025930: 12 February 2010

Expiry date of the certificate: 11 February 2015

Test report: 936/21211670/A of 29 October 2009,
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

Publication: BAnz. 12 February 2010, no. 24, p. 552, chapter I, no. 1.2,
UBA announcement of 25 January 2010

Supplementary testing according to EN 15267:

Certificate No. 0000025930_01: 2 August 2010

Expiry date of the certificate: 11 February 2015

Test report: 936/21211670/B of 26 March 2010,
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

Publication: BAnz. 28 July 2010, no. 111, p. 2597, chapter I, no. 2.1,
UBA announcement of 12 July 2010

Supplementary testing according to EN 15267:

Certificate No. 0000025930_02: 16 March 2012

Expiry date of the certificate: 11 February 2015

Test report: 936/21217568/A of 18 October 2011,
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz. 2 March 2012, no. 36, p. 920, chapter I, no. 5.1,
UBA announcement of 23 February 2012

Notifications

Statements of TÜV Rheinland Energie und Umwelt GmbH of 24 September 2010, 5 and 8 November 2010

Publication: BAnz. 26 January 2011, no. 14, p. 294, chapter IV, notifications 2 and 30 (sequential no. 13) (new manufacturer name, new software version)

UBA announcement of 10 January 2011

Statement of TÜV Rheinland Energie und Umwelt GmbH of 25 March 2013

Publication: BAnz AT 23 July 2013 B4, chapter V, notifications 12 (sequential no. 12) and 13 (sequential no. 13) (new manufacturer name, new software version)

UBA announcement of 3 July 2013

Statement of TÜV Rheinland Energie und Umwelt GmbH of 2 April 2014

Publication: BAnz AT 5 August 2014 B11, chapter V, notification 13 (new software version, chopper motor)

UBA announcement of 17 July 2014

Renewal of the certificate

Certificate No. 0000025930_03: 2 February 2015

Expiry date of the certificate: 11 February 2020

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for CO
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland
	2011-10-18

Measured component

	CO
Certification range	0 - 75 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	1.80 mg/m ³
Sum of negative CS at zero point	-1.30 mg/m ³
Sum of positive CS at reference point	1.07 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	1.80 mg/m ³
Uncertainty of cross sensitivity	1.039 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.747 mg/m ³	0.558 (mg/m ³) ²
Lack of fit	u _{lof} 0.289 mg/m ³	0.084 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.346 mg/m ³	0.120 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.866 mg/m ³	0.750 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.751 mg/m ³	0.564 (mg/m ³) ²
Influence of supply voltage	u _v 0.115 mg/m ³	0.013 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.039 mg/m ³	1.080 (mg/m ³) ²
Influence of sample gas flow	u _b -0.029 mg/m ³	0.001 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	1.88 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	3.69 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m ³	7.4
U in % of the ELV 50 mg/m ³	10.0
U in % of the ELV 50 mg/m ³	7.5

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 MULTOR for CO
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland
	2011-10-18

Measured component

	CO
Certification range	0 - 200 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	6.76 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	6.76 mg/m ³
Uncertainty of cross sensitivity	3.903 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.588 mg/m ³	2.522 (mg/m ³) ²
Lack of fit	u _{lof} 1.155 mg/m ³	1.334 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.924 mg/m ³	0.854 (mg/m ³) ²
Span drift from field test	u _{d,s} -3.002 mg/m ³	9.012 (mg/m ³) ²
Influence of ambient temperature at span	u _t 2.406 mg/m ³	5.789 (mg/m ³) ²
Influence of supply voltage	u _v 0.157 mg/m ³	0.025 (mg/m ³) ²
Cross sensitivity (interference)	u _i 3.903 mg/m ³	15.233 (mg/m ³) ²
Influence of sample gas flow	u _b 0.127 mg/m ³	0.016 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 1.617 mg/m ³	2.613 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{\max j})^2} \quad 6.12 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 11.99 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 160 mg/m³ 7.5

U in % of the ELV 160 mg/m³ 10.0

U in % of the ELV 160 mg/m³ 7.5

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for NO
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland 2011-10-18

Measured component

	NO
Certification range	0 - 100 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	1.56 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	2.46 mg/m ³
Sum of negative CS at reference point	-0.73 mg/m ³
Maximum sum of cross sensitivities	2.46 mg/m ³
Uncertainty of cross sensitivity	1.420 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.191 mg/m ³	1.418 (mg/m ³) ²
Lack of fit	u _{lof} 0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} -1.212 mg/m ³	1.469 (mg/m ³) ²
Span drift from field test	u _{d,s} 1.732 mg/m ³	3.000 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.529 mg/m ³	0.280 (mg/m ³) ²
Influence of supply voltage	u _v 0.142 mg/m ³	0.020 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.420 mg/m ³	2.017 (mg/m ³) ²
Influence of sample gas flow	u _b -0.104 mg/m ³	0.011 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.808 mg/m ³	0.653 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 2.99 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 5.85 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 11.7

U in % of the ELV 50 mg/m³ 20.0

U in % of the ELV 50 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 MULTOR for NO
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland 2011-10-18

Measured component

Certification range	NO 0 - 250 mg/m ³
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	8.95 mg/m ³
Sum of negative CS at zero point	-4.43 mg/m ³
Sum of positive CS at reference point	3.45 mg/m ³
Sum of negative CS at reference point	-3.65 mg/m ³
Maximum sum of cross sensitivities	8.95 mg/m ³
Uncertainty of cross sensitivity	5.167 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 2.241 mg/m ³	5.022 (mg/m ³) ²
Lack of fit	u _{lof} -1.155 mg/m ³	1.334 (mg/m ³) ²
Zero drift from field test	u _{d,z} 2.742 mg/m ³	7.519 (mg/m ³) ²
Span drift from field test	u _{d,s} 4.186 mg/m ³	17.523 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.950 mg/m ³	0.903 (mg/m ³) ²
Influence of supply voltage	u _v 0.737 mg/m ³	0.543 (mg/m ³) ²
Cross sensitivity (interference)	u _i 5.167 mg/m ³	26.701 (mg/m ³) ²
Influence of sample gas flow	u _b 0.277 mg/m ³	0.077 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 2.021 mg/m ³	4.083 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 7.98 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 15.64 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 131 mg/m³ 11.9

U in % of the ELV 131 mg/m³ 20.0

U in % of the ELV 131 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 DEFOR for NO
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	UVRAS

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland 2011-10-18

Measured component

	NO
Certification range	0 - 50 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	1.86 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	1.06 mg/m ³
Sum of negative CS at reference point	-0.94 mg/m ³
Maximum sum of cross sensitivities	1.86 mg/m ³
Uncertainty of cross sensitivity	1.074 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.751 mg/m ³	0.564 (mg/m ³) ²
Lack of fit	u _{lof} -0.115 mg/m ³	0.013 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.375 mg/m ³	0.141 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.866 mg/m ³	0.750 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.153 mg/m ³	0.023 (mg/m ³) ²
Influence of supply voltage	u _v 0.233 mg/m ³	0.054 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.074 mg/m ³	1.153 (mg/m ³) ²
Influence of sample gas flow	u _b 0.052 mg/m ³	0.003 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 1.69 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 3.32 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 30 mg/m³ 11.1

U in % of the ELV 30 mg/m³ 20.0

U in % of the ELV 30 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 DEFOR for NO ₂
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	UVRAS

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

	NO ₂
Certification range	0 - 50 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	1.72 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	1.93 mg/m ³
Sum of negative CS at reference point	-0.26 mg/m ³
Maximum sum of cross sensitivities	1.93 mg/m ³
Uncertainty of cross sensitivity	1.114 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.520 mg/m ³	0.270 (mg/m ³) ²
Lack of fit	u _{lof} -0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.693 mg/m ³	0.480 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.866 mg/m ³	0.750 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.458 mg/m ³	0.210 (mg/m ³) ²
Influence of supply voltage	u _v 0.110 mg/m ³	0.012 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.114 mg/m ³	1.242 (mg/m ³) ²
Influence of sample gas flow	u _b 0.030 mg/m ³	0.001 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	1.78 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	3.50 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m ³	7.0
U in % of the ELV 50 mg/m ³	20.0
U in % of the ELV 50 mg/m ³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for SO ₂
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	SO ₂ 0 - 75 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	2.75 mg/m ³
Sum of negative CS at zero point	-1.75 mg/m ³
Sum of positive CS at reference point	2.30 mg/m ³
Sum of negative CS at reference point	-1.82 mg/m ³
Maximum sum of cross sensitivities	2.75 mg/m ³
Uncertainty of cross sensitivity	1.585 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.228 mg/m ³	1.508 (mg/m ³) ²
Lack of fit	u _{lof} 0.410 mg/m ³	0.168 (mg/m ³) ²
Zero drift from field test	u _{d,z} -1.212 mg/m ³	1.469 (mg/m ³) ²
Span drift from field test	u _{d,s} 1.299 mg/m ³	1.687 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.929 mg/m ³	0.863 (mg/m ³) ²
Influence of supply voltage	u _v 0.227 mg/m ³	0.052 (mg/m ³) ²
Cross sensitivity (interference)	u _i 1.585 mg/m ³	2.512 (mg/m ³) ²
Influence of sample gas flow	u _b 0.057 mg/m ³	0.003 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 2.94 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 5.76 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 11.5

U in % of the ELV 50 mg/m³ 20.0

U in % of the ELV 50 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 MULTOR for SO ₂
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	SO ₂ 0 - 250 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	9.63 mg/m ³
Sum of negative CS at zero point	-2.65 mg/m ³
Sum of positive CS at reference point	5.93 mg/m ³
Sum of negative CS at reference point	-1.20 mg/m ³
Maximum sum of cross sensitivities	9.63 mg/m ³
Uncertainty of cross sensitivity	5.557 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.546 mg/m ³	2.390 (mg/m ³) ²
Lack of fit	u _{lof} -2.714 mg/m ³	7.366 (mg/m ³) ²
Zero drift from field test	u _{d,z} 2.115 mg/m ³	4.473 (mg/m ³) ²
Span drift from field test	u _{d,s} -3.002 mg/m ³	9.012 (mg/m ³) ²
Influence of ambient temperature at span	u _t 2.901 mg/m ³	8.416 (mg/m ³) ²
Influence of supply voltage	u _v 0.839 mg/m ³	0.704 (mg/m ³) ²
Cross sensitivity (interference)	u _i 5.557 mg/m ³	30.880 (mg/m ³) ²
Influence of sample gas flow	u _b -0.410 mg/m ³	0.168 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 2.021 mg/m ³	4.083 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	8.22 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	16.10 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 150 mg/m ³	10.7
U in % of the ELV 150 mg/m ³	20.0
U in % of the ELV 150 mg/m ³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 DEFOR for SO ₂
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	UVRAS

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	SO ₂ 0 - 75 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	-0.81 mg/m ³
Sum of positive CS at reference point	0.35 mg/m ³
Sum of negative CS at reference point	-2.91 mg/m ³
Maximum sum of cross sensitivities	-2.91 mg/m ³
Uncertainty of cross sensitivity	-1.680 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 1.206 mg/m ³	1.454 (mg/m ³) ²
Lack of fit	u _{lof} -0.404 mg/m ³	0.163 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.606 mg/m ³	0.367 (mg/m ³) ²
Span drift from field test	u _{d,s} 1.299 mg/m ³	1.687 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.964 mg/m ³	0.929 (mg/m ³) ²
Influence of supply voltage	u _v 0.067 mg/m ³	0.004 (mg/m ³) ²
Cross sensitivity (interference)	u _i -1.680 mg/m ³	2.823 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.606 mg/m ³	0.368 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 2.79 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 5.47 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 50 mg/m³ 10.9

U in % of the ELV 50 mg/m³ 20.0

U in % of the ELV 50 mg/m³ 15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for CH ₄
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	CH ₄ 0 - 50 mg/m ³
---------------------	---

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.48 mg/m ³
Sum of negative CS at zero point	-1.77 mg/m ³
Sum of positive CS at reference point	0.00 mg/m ³
Sum of negative CS at reference point	-0.63 mg/m ³
Maximum sum of cross sensitivities	-1.77 mg/m ³
Uncertainty of cross sensitivity	-1.022 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.630 mg/m ³	0.397 (mg/m ³) ²
Lack of fit	u _{lof} 0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} 0.520 mg/m ³	0.270 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.635 mg/m ³	0.403 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.416 mg/m ³	0.173 (mg/m ³) ²
Influence of supply voltage	u _v 0.306 mg/m ³	0.094 (mg/m ³) ²
Cross sensitivity (interference)	u _i -1.022 mg/m ³	1.044 (mg/m ³) ²
Influence of sample gas flow	u _p -0.035 mg/m ³	0.001 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{\max, j})^2} \quad 1.61 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 3.16 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 20 mg/m³ **15.8**

Requirement of 2000/76/EC and 2001/80/EC

U in % of the ELV 20 mg/m³ **30.0 ****

Requirement of EN 15267-3

U in % of the ELV 20 mg/m³ **22.5**

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 30 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak
Name of measuring system	MAC GMS800 MULTOR for CH ₄
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	CH ₄ 0 - 286 mg/m ³
---------------------	--

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	1.06 mg/m ³
Sum of negative CS at reference point	-1.49 mg/m ³
Maximum sum of cross sensitivities	-1.49 mg/m ³
Uncertainty of cross sensitivity	-0.859 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Repeatability standard deviation at set point *	u _r 0.620 mg/m ³	0.384 (mg/m ³) ²
Lack of fit	u _{lof} -1.501 mg/m ³	2.253 (mg/m ³) ²
Zero drift from field test	u _{d,z} 1.156 mg/m ³	1.336 (mg/m ³) ²
Span drift from field test	u _{d,s} -2.972 mg/m ³	8.833 (mg/m ³) ²
Influence of ambient temperature at span	u _t 2.843 mg/m ³	8.083 (mg/m ³) ²
Influence of supply voltage	u _v 0.532 mg/m ³	0.283 (mg/m ³) ²
Cross sensitivity (interference)	u _i -0.859 mg/m ³	0.737 (mg/m ³) ²
Influence of sample gas flow	u _p 0.370 mg/m ³	0.137 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 2.312 mg/m ³	5.344 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	5.23 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	10.26 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 100 mg/m ³	10.3
U in % of the ELV 100 mg/m ³	30.0 **
U in % of the ELV 100 mg/m ³	22.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 30 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for N ₂ O
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

	N ₂ O
Certification range	0 - 50 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.93 mg/m ³
Sum of negative CS at zero point	-1.41 mg/m ³
Sum of positive CS at reference point	0.00 mg/m ³
Sum of negative CS at reference point	-0.65 mg/m ³
Maximum sum of cross sensitivities	-1.41 mg/m ³
Uncertainty of cross sensitivity	-0.814 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.410 mg/m ³	0.168 (mg/m ³) ²
Lack of fit	u _{lof} 0.231 mg/m ³	0.053 (mg/m ³) ²
Zero drift from field test	u _{d,z} -0.318 mg/m ³	0.101 (mg/m ³) ²
Span drift from field test	u _{d,s} 0.866 mg/m ³	0.750 (mg/m ³) ²
Influence of ambient temperature at span	u _t 0.436 mg/m ³	0.190 (mg/m ³) ²
Influence of supply voltage	u _v 0.172 mg/m ³	0.030 (mg/m ³) ²
Cross sensitivity (interference)	u _i -0.814 mg/m ³	0.663 (mg/m ³) ²
Influence of sample gas flow	u _p 0.052 mg/m ³	0.003 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.404 mg/m ³	0.163 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	1.46 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	2.85 mg/m ³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the range 50 mg/m ³	5.7
U in % of the range 50 mg/m ³	20.0 **
U in % of the range 50 mg/m ³	15.0

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 20 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 UNOR for CO ₂
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	NDIR

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland
	2011-10-18

Measured component

Certification range	CO ₂	0 - 25 Vol.-%
---------------------	-----------------	---------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00	Vol.-%
Sum of negative CS at zero point	-0.47	Vol.-%
Sum of positive CS at reference point	0.00	Vol.-%
Sum of negative CS at reference point	0.00	Vol.-%
Maximum sum of cross sensitivities	-0.47	Vol.-%
Uncertainty of cross sensitivity	-0.271	Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

		u	u ²
Standard deviation from paired measurements under field conditions *	u _D	0.156 Vol.-%	0.024 (Vol.-%) ²
Lack of fit	u _{lof}	-0.144 Vol.-%	0.021 (Vol.-%) ²
Zero drift from field test	u _{d,z}	-0.188 Vol.-%	0.035 (Vol.-%) ²
Span drift from field test	u _{d,s}	0.346 Vol.-%	0.120 (Vol.-%) ²
Influence of ambient temperature at span	u _t	0.300 Vol.-%	0.090 (Vol.-%) ²
Influence of supply voltage	u _v	0.049 Vol.-%	0.002 (Vol.-%) ²
Cross sensitivity (interference)	u _i	-0.271 Vol.-%	0.074 (Vol.-%) ²
Influence of sample gas flow	u _p	0.017 Vol.-%	0.000 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	0.202 Vol.-%	0.041 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.64 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.25 Vol.-%

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC	U in % of the ELV 25 Vol.-%	5.0
Requirement of EN 15267-3	U in % of the ELV 25 Vol.-%	10.0 **
	U in % of the ELV 25 Vol.-%	7.5

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.
A value of 10 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 OXOR-P for O ₂
Serial number of the candidates	TÜV 1 / TÜV 3
Measuring principle	paramagnetic

Test report

Test laboratory	936/21217568/A TÜV Rheinland
Date of report	2011-10-18

Measured component

Certification range	O ₂ 0 - 25 Vol.-%
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.00 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.00 Vol.-%
Uncertainty of cross sensitivity	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.084 Vol.-%	0.007 (Vol.-%) ²
Lack of fit	u _{lof} -0.040 Vol.-%	0.002 (Vol.-%) ²
Zero drift from field test	u _{d,z} 0.120 Vol.-%	0.014 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.120 Vol.-%	0.014 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.110 Vol.-%	0.012 (Vol.-%) ²
Influence of supply voltage	u _v 0.003 Vol.-%	0.000 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.000 Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _p -0.023 Vol.-%	0.001 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.202 Vol.-%	0.041 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{\max, j})^2} \quad 0.30 \text{ Vol.-%}$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 0.59 \text{ Vol.-%}$$

Relative total expanded uncertainty

U in % of the range 25 Vol.-% **2.4**

Requirement of 2000/76/EC and 2001/80/EC

U in % of the range 25 Vol.-% **10.0 ****

Requirement of EN 15267-3

U in % of the range 25 Vol.-% **7.5**

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 10 % was used for this.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick Maihak GmbH
Name of measuring system	MAC GMS800 OXOR-E for O ₂
Serial number of the candidates	TÜV 2 / TÜV 4
Measuring principle	electrochemical cell

Test report

Test laboratory	936/21217568/A
Date of report	TÜV Rheinland 2011-10-18

Measured component

Certification range	O ₂ 0 - 25 Vol.-%
---------------------	---------------------------------

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at reference point	0.33 Vol.-%
Sum of negative CS at reference point	0.00 Vol.-%
Maximum sum of cross sensitivities	0.33 Vol.-%
Uncertainty of cross sensitivity	0.191 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

	u	u ²
Standard deviation from paired measurements under field conditions *	u _D 0.108 Vol.-%	0.012 (Vol.-%) ²
Lack of fit	u _{lof} 0.058 Vol.-%	0.003 (Vol.-%) ²
Zero drift from field test	u _{d,z} 0.120 Vol.-%	0.014 (Vol.-%) ²
Span drift from field test	u _{d,s} 0.120 Vol.-%	0.014 (Vol.-%) ²
Influence of ambient temperature at span	u _t 0.127 Vol.-%	0.016 (Vol.-%) ²
Influence of supply voltage	u _v 0.030 Vol.-%	0.001 (Vol.-%) ²
Cross sensitivity (interference)	u _i 0.191 Vol.-%	0.036 (Vol.-%) ²
Influence of sample gas flow	u _p 0.029 Vol.-%	0.001 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.202 Vol.-%	0.041 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{\max, j})^2} \quad 0.37 \text{ Vol.-%}$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 0.73 \text{ Vol.-%}$$

Relative total expanded uncertainty

U in % of the range 25 Vol.-% **2.9**

Requirement of 2000/76/EC and 2001/80/EC

U in % of the range 25 Vol.-% **10.0 ****

Requirement of EN 15267-3

U in % of the range 25 Vol.-% **7.5**

** For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given.

A value of 10 % was used for this.