

# CONFIRMATION

Notification: 0000025930\_04\_06  
on changes according to EN 15267 regarding certificate 0000025930\_04 dated 12 February 2020

---

**AMS designation:** PowerCEMS100 for CO, NO, NO<sub>2</sub>, N<sub>2</sub>O, SO<sub>2</sub>, CH<sub>4</sub>, O<sub>2</sub> and CO<sub>2</sub>

**Manufacturer:** SICK AG  
Gisela Sick Straße 1  
79276 Reute  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 31 August 2024  
Publication: BAnz AT 31.10.2024 B9**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**40 Notification as regards Federal Environment Agency (UBA) notices  
of 23 February 2012 (BAnz. p. 920, chapter I number 5.1) and  
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 44)**

The current software versions for the Power-CEMS100 modular measuring system for  
CO, NO, NO<sub>2</sub>, SO<sub>2</sub>, CH<sub>4</sub>, N<sub>2</sub>O, CO<sub>2</sub> and O<sub>2</sub> from the company SICK AG are:


BCU: 9150883\_4.006  
UNOR/MULTOR: 9137995\_4.002  
DEFOR: 9139736\_4.004  
OXOR: 9138052\_4.001  
Gas module: 9134803\_4.003


In future, the DEFOR module can also be equipped with alternative interference filters  
for NO<sub>2</sub> measurement with the article numbers 5347371 and 5347372.

Statement issued by TÜV Rheinland Energy & Environment GmbH  
dated 04 March 2024

---

TÜV Rheinland Energy & Environment GmbH  
Cologne, 12 November 2024

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.