

CONFIRMATION

Notification: 0000035015_04_01
on changes according to EN 15267 regarding certificate 0000035015_04 dated 16 February 2022

AMS designation: MERCEM 300Z / indoor for Hg

Manufacturer: SICK AG
Gisela Sick Straße 1
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**27 Notification as regards Federal Environment Agency (UBA) notices
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 2.3) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 55)**


The current software version for the measuring devices MERCEM300Z and
MERCEM300Z Indoor for Hg from SICK AG is now
9191789_17EF.

Furthermore, the measuring device can be equipped with a revised lamp assembly with
the material number 2108573.

Statement issued by TÜV Rheinland Energy GmbH dated 15 February 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.