

# CONFIRMATION

Notification: 0000035015\_04\_05  
on changes according to EN 15267 regarding certificate 0000035015\_04 dated 16 February 2022

---

**AMS designation:** MERCEM 300Z / indoor for Hg

**Manufacturer:** Endress+Hauser SICK GmbH + Co. KG  
Bergener Ring 27  
01458 Ottendorf-Okrilla  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 27 August 2025  
Publication: BAnz AT 31.10.2025 B5**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

- 11 Notification as regards Federal Environment Agency (UBA) notices  
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 2.3) and  
of 2 April 2025 (BAnz AT 19.05.2025 B3, chapter IV notification 21)**


The standard version of the MERCEM300Z measuring device for Hg  
from Endress+Hauser SICK GmbH+Co. KG


can now also be installed in the Rittal equipment cabinet with the model number  
CS9868.500. To do this, the enclosure must be equipped with the SK 3185.330 climate  
control unit and the SK 3105.390 enclosure heater with fan, both also from Rittal. This  
change does not apply to the "indoor" version of the measuring device.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 May 2025

---

TÜV Rheinland Energy & Environment GmbH  
Cologne, 21 November 2025

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
qa11-info@tuv.com  
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.