

CONFIRMATION

Notification: 0000062068_01_05
on changes according to EN 15267 regarding certificate 0000062068_01 dated 5 November 2019

AMS designation: MCS200HW for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, TOC, O₂, H₂O and CO₂

Manufacturer: SICK AG
Gisela Sick Straße 1
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**38 Notification as regards Federal Environment Agency (UBA) notices
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter I number 1.4) and
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 39)**

The current software versions for the MCS200HW modular measuring system
for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, H₂O, CO₂, TOC and O₂
from SICK AG are as follows:


MCS200HW: 9264565_1.7.7.4R_1C6E


GMS811 FIDORi: Firmware: 9230690_4.003

BCU: 9150883_4.006

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.