

CONFIRMATION

Notification: 0000062068_02_01
on changes according to EN 15267 regarding certificate 0000062068_02 dated 3 July 2024

AMS designation: MCS200HW for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, TOC, O₂, H₂O and CO₂

Manufacturer: SICK AG
Gisela Sick Straße 1
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 31 August 2024
Publication: BAnz AT 31.10.2024 B9**


**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


- 44 Notification as regards Federal Environment Agency (UBA) notices
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter I number 1.4) and
of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter V notification 38)**

The MCS200HW emission measuring system for CO, NO, NO₂, N₂O, SO₂, HCl, NH₃, CH₄, H₂O, CO₂, O₂ and TOC from the company SICK AG is also suitable for use in installations subject to the 44th BImSchV.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 10 May 2024

TÜV Rheinland Energy & Environment GmbH
Cologne, 12 November 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.