

CONFIRMATION

Notification: 0000062064_01_02
on changes according to EN 15267 regarding certificate 0000062064_01 dated 20 March 2024

AMS designation: 49iQ for O₃

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 27 August 2025
Publication: BAnz AT 31.10.2025 B5**


**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

- 43 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter III number 3.1) and
of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter IV notification 66)**


The 49iQ ambient air measuring device for O₃ from Thermo Fisher Scientific can be optionally equipped with a NAFION(TM) sample gas dryer. In addition, the ozone scrubber will also be replaced. In future, the scrubber (type no. 14697 (drawing # 45P229)), which is already used in the suitability-tested Thermo 49i ozone analyser, will also be used in the 49iQ model.

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 30 January 2025

TÜV Rheinland Energy & Environment GmbH
Cologne, 21 November 2025



i. V. Dipl.-Ing. Guido Baum



i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.