

CONFIRMATION

Notification: 0000062067_01_01
on changes according to EN 15267 regarding certificate 0000062067_01 dated 20 March 2024

AMS designation: 42iQ for NO, NO₂ and NO_x

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**63 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2019 (BAnz AT 26.03.2019, chapter III number 1.1) and
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 81)**


The current software version for the 42iQ ambient air quality measuring system
for NO, NO₂ and NO_x from Thermo Fisher Scientific is: 01.06.15.34700

The measuring system can be operated with the modified gas flow scheme for the waste
gas flow of the permeation dryer.

Statement issued by TÜV Rheinland Energy GmbH dated 15 December 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.